

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082324\
 Data File : PP066855.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2024 00:44
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/26/2024
 Supervised By :Ankita Jodhani 08/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 04:11:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 24 04:04:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.676	3.965	5875579	4581457	4.180	4.166
2) SA Decachlor...	10.581	9.144	3435973	2665704	4.153	3.855
Target Compounds						
26) L6 AR-1254-1	6.705	5.890	2037110	2566274	45.160	48.774
27) L6 AR-1254-2	6.924	6.039	2893062	2201216	43.845	47.647
28) L6 AR-1254-3	7.291	6.452	2576252	2994945	40.104	45.338
29) L6 AR-1254-4	7.576	6.682	1568234	1544084	39.292	47.153
30) L6 AR-1254-5	7.994	7.108	2526377	2173148	53.406m	42.822

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082324\
Data File : PP066855.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2024 00:44
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/26/2024
Supervised By :Ankita Jodhani 08/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 04:11:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 24 04:04:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

