

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082421\
 Data File : PP038843.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Aug 2021 14:47
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 02:17:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.885	3.867	2051871	1336429	50.842	55.768
2)	SA Decachlor...	10.870	9.127	1329889	1187316	54.052	48.749
Target Compounds							
16)	L4 AR-1242-1	6.203	5.111	464790	346034	466.625	501.180
17)	L4 AR-1242-2	6.228	5.130	664385	495474	462.345	507.569
18)	L4 AR-1242-3	6.293	5.320	420778	266809	470.013	507.382
19)	L4 AR-1242-4	6.400	5.418	344290	245520	470.185	513.637
20)	L4 AR-1242-5	7.183	5.979	319128	310912	446.359	502.242



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082421\
Data File : PP038843.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Aug 2021 14:47
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 02:17:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.885 min
Delta R.T.: -0.003 min
Response: 2051871
Conc: 50.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.867 min
Delta R.T.: -0.007 min
Response: 1336429
Conc: 55.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.870 min
Delta R.T.: -0.007 min
Response: 1329889
Conc: 54.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.127 min
Delta R.T.: -0.009 min
Response: 1187316
Conc: 48.75 ng/ml



#16 AR-1242-1


R.T.: 6.203 min
Delta R.T.: -0.002 min
Response: 464790
Conc: 466.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.111 min
Delta R.T.: -0.006 min
Response: 346034
Conc: 501.18 ng/ml




#17 AR-1242-2

R.T.: 6.228 min
Delta R.T.: -0.001 min
Response: 664385
Conc: 462.35 ng/ml



#17 AR-1242-2


R.T.: 5.130 min
Delta R.T.: -0.006 min
Response: 495474
Conc: 507.57 ng/ml



#18 AR-1242-3


R.T.: 6.293 min
Delta R.T.: -0.002 min
Response: 420778
Conc: 470.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.320 min
Delta R.T.: -0.006 min
Response: 266809
Conc: 507.38 ng/ml




#19 AR-1242-4

R.T.: 6.400 min
Delta R.T.: -0.002 min
Response: 344290
Conc: 470.18 ng/ml



#19 AR-1242-4


R.T.: 5.418 min
Delta R.T.: -0.006 min
Response: 245520
Conc: 513.64 ng/ml



#20 AR-1242-5

R.T.: 7.183 min
Delta R.T.: -0.002 min
Response: 319128
Conc: 446.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.979 min
Delta R.T.: -0.006 min
Response: 310912
Conc: 502.24 ng/ml