

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082421\
 Data File : PP038876.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Aug 2021 1:13
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 09:52:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.880	3.866	2020388	1325147	50.062	55.297
2)	SA Decachlor...	10.860	9.121	1352268	1298325	54.962	53.307
Target Compounds							
26)	L6 AR-1254-1	7.107	5.975	540870	653358	474.023	497.499
27)	L6 AR-1254-2	7.336	6.136	752911	587347	436.001	511.727
28)	L6 AR-1254-3	7.718	6.552	796772	941513	443.854	505.985
29)	L6 AR-1254-4	8.014	6.792	571081	594740	442.590	508.921
30)	L6 AR-1254-5	8.443	7.219	587546	807141	475.049	499.372



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082421\
Data File : PP038876.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Aug 2021 1:13
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 09:52:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.880 min
Delta R.T.: -0.008 min
Response: 2020388
Conc: 50.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.866 min
Delta R.T.: -0.008 min
Response: 1325147
Conc: 55.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.860 min
Delta R.T.: -0.010 min
Response: 1352268
Conc: 54.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.121 min
Delta R.T.: -0.015 min
Response: 1298325
Conc: 53.31 ng/ml



#26 AR-1254-1


R.T.: 7.107 min
Delta R.T.: -0.006 min
Response: 540870
Conc: 474.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.975 min
Delta R.T.: -0.008 min
Response: 653358
Conc: 497.50 ng/ml




#27 AR-1254-2

R.T.: 7.336 min
Delta R.T.: -0.007 min
Response: 752911
Conc: 436.00 ng/ml



#27 AR-1254-2


R.T.: 6.136 min
Delta R.T.: -0.008 min
Response: 587347
Conc: 511.73 ng/ml



#28 AR-1254-3


R.T.: 7.718 min
Delta R.T.: -0.007 min
Response: 796772
Conc: 443.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.552 min
Delta R.T.: -0.008 min
Response: 941513
Conc: 505.98 ng/ml




#29 AR-1254-4

R.T.: 8.014 min
Delta R.T.: -0.008 min
Response: 571081
Conc: 442.59 ng/ml



#29 AR-1254-4


R.T.: 6.792 min
Delta R.T.: -0.008 min
Response: 594740
Conc: 508.92 ng/ml



#30 AR-1254-5

R.T.: 8.443 min
Delta R.T.: -0.008 min
Response: 587546
Conc: 475.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.219 min
Delta R.T.: -0.008 min
Response: 807141
Conc: 499.37 ng/ml