

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 07:55:40 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038905.D 1000 =PP038903.D 750 =PP038904.D
 250 =PP038906.D 50 =PP038907.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	3.964	3.748	3.884	3.992	4.302	3.978	E4	5.14
2)	SA	Decachlorobiphenyl	2.508	2.363	2.475	2.547	2.740	2.526	E4	5.45
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1	0.928	0.824	0.869	0.981	1.109	0.942	E3	11.72
17)	L4	AR-1242-2	1.308	1.184	1.254	1.393	1.562	1.340	E3	10.86
18)	L4	AR-1242-3	8.241	7.342	7.793	8.743	9.386	8.301	E2	9.63
19)	L4	AR-1242-4	6.765	6.088	6.447	6.832	7.821	6.790	E2	9.53
20)	L4	AR-1242-5	6.164	5.450	5.845	6.412	7.847	6.344	E2	14.42
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP038905.D 1000 =PP038903.D 750 =PP038904.D
 250 =PP038906.D 50 =PP038907.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.544	2.471	2.530	2.491	2.440	2.495	E4	1.70
2)	SA	Decachlorobiphenyl	2.440	2.220	2.325	2.536	2.754	2.455	E4	8.36
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 07:55:40 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038905.D 1000 =PP038903.D 750 =PP038904.D
 250 =PP038906.D 50 =PP038907.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	6.663	6.102	6.391	6.675	7.086	6.584 E2	5.56
17) L4	AR-1242-2	9.361	8.739	9.117	9.493	9.473	9.237 E2	3.42
18) L4	AR-1242-3	5.008	4.702	4.882	4.856	4.856	4.861 E2	2.24
19) L4	AR-1242-4	4.569	4.265	4.427	4.588	4.678	4.506 E2	3.59
20) L4	AR-1242-5	5.917	5.530	5.746	5.610	5.702	5.701 E2	2.58
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range