

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 08:17:44 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038910.D 1000 =PP038908.D 750 =PP038909.D
 250 =PP038911.D 50 =PP038912.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.878	3.686	3.728	3.929	4.075	3.859	E4	4.07
2) SA	Decachlorobiphenyl	2.474	2.372	2.386	2.535	2.665	2.486	E4	4.82
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	7.050	6.351	6.546	7.537	7.876	7.072	E2	9.11
22) L5	AR-1248-2	0.856	0.759	0.788	0.912	1.006	0.864	E3	11.48
23) L5	AR-1248-3	0.987	0.887	0.916	1.040	1.021	0.970	E3	6.83
24) L5	AR-1248-4	1.042	0.936	0.967	1.108	1.201	1.051	E3	10.23
25) L5	AR-1248-5	1.031	0.925	0.956	1.095	1.161	1.033	E3	9.43
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP038910.D 1000 =PP038908.D 750 =PP038909.D
 250 =PP038911.D 50 =PP038912.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.504	2.440	2.441	2.443	2.276	2.421	E4	3.53
2) SA	Decachlorobiphenyl	2.428	2.242	2.292	2.484	2.534	2.396	E4	5.20
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 08:17:44 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038910.D 1000 =PP038908.D 750 =PP038909.D
 250 =PP038911.D 50 =PP038912.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	4.908	4.550	4.639	4.800	4.819	4.743 E2	3.07
22) L5	AR-1248-2	6.654	6.091	6.252	6.698	6.892	6.517 E2	5.11
23) L5	AR-1248-3	6.913	6.382	6.540	6.972	6.999	6.761 E2	4.16
24) L5	AR-1248-4	8.368	7.779	7.924	8.316	8.183	8.114 E2	3.13
25) L5	AR-1248-5	8.539	8.063	8.166	8.567	7.791	8.225 E2	4.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range