

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 08:33:08 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038915.D 1000 =PP038913.D 750 =PP038914.D
 250 =PP038916.D 50 =PP038917.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.925	3.785	3.840	4.045	4.394	3.998	E4	6.06
2) SA	Decachlorobiphenyl	2.615	2.484	2.495	2.632	2.795	2.604	E4	4.85
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	0.990	0.904	0.937	1.094	1.292	1.043	E3	15.02
27) L6	AR-1254-2	1.456	1.328	1.379	1.580	1.828	1.514	E3	13.17
28) L6	AR-1254-3	1.486	1.363	1.417	1.614	1.824	1.541	E3	11.94
29) L6	AR-1254-4	1.080	0.994	1.032	1.163	1.332	1.120	E3	12.00
30) L6	AR-1254-5	1.101	1.023	1.054	1.182	1.285	1.129	E3	9.35
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP038915.D 1000 =PP038913.D 750 =PP038914.D
 250 =PP038916.D 50 =PP038917.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.523	2.491	2.514	2.509	2.417	2.491	E4	1.71
2) SA	Decachlorobiphenyl	2.473	2.293	2.369	2.585	2.778	2.500	E4	7.62
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082521.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 26 08:33:08 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038915.D 1000 =PP038913.D 750 =PP038914.D
 250 =PP038916.D 50 =PP038917.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.227	1.156	1.190	1.282	1.305	1.232 E3	5.03
27) L6	AR-1254-2	1.077	0.994	1.031	1.137	1.198	1.087 E3	7.52
28) L6	AR-1254-3	1.751	1.662	1.699	1.806	1.767	1.737 E3	3.27
29) L6	AR-1254-4	1.123	1.064	1.089	1.151	1.123	1.110 E3	3.04
30) L6	AR-1254-5	1.548	1.473	1.511	1.574	1.476	1.516 E3	2.92
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range