

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082521\
 Data File : PP038909.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Aug 2021 21:23
 Operator : AJ\MA
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 08:12:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 26 08:11:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.877	3.861	2796311	1830698	74.292	74.368
2)	SA Decachlor...	10.857	9.119	1789128	1718734	74.219	74.058
Target Compounds							
21)	L5 AR-1248-1	6.196	5.105	490985	347891	738.402	740.397
22)	L5 AR-1248-2	6.490	5.368	590834	468911	737.829	740.473
23)	L5 AR-1248-3	6.707	5.412	687269	490489	738.903	741.860
24)	L5 AR-1248-4	7.136	5.594	725263	594324	738.732	740.712
25)	L5 AR-1248-5	7.175	6.015	716758	612430	738.629	741.823



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082521\
Data File : PP038909.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Aug 2021 21:23
Operator : AJ\MA
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 08:12:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 26 08:11:52 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 2796311
Conc: 74.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.861 min
Delta R.T.: 0.000 min
Response: 1830698
Conc: 74.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.857 min
Delta R.T.: 0.000 min
Response: 1789128
Conc: 74.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.119 min
Delta R.T.: 0.000 min
Response: 1718734
Conc: 74.06 ng/ml



#21 AR-1248-1


R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 490985
Conc: 738.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 5.105 min
Delta R.T.: 0.000 min
Response: 347891
Conc: 740.40 ng/ml




#22 AR-1248-2

R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 590834
Conc: 737.83 ng/ml



#22 AR-1248-2


R.T.: 5.368 min
Delta R.T.: 0.000 min
Response: 468911
Conc: 740.47 ng/ml



#23 AR-1248-3


R.T.: 6.707 min
Delta R.T.: 0.000 min
Response: 687269
Conc: 738.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.412 min
Delta R.T.: 0.000 min
Response: 490489
Conc: 741.86 ng/ml




#24 AR-1248-4

R.T.: 7.136 min
Delta R.T.: 0.000 min
Response: 725263
Conc: 738.73 ng/ml



#24 AR-1248-4


R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 594324
Conc: 740.71 ng/ml



#25 AR-1248-5

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 716758
Conc: 738.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 6.015 min
Delta R.T.: 0.000 min
Response: 612430
Conc: 741.82 ng/ml