

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082523\
 Data File : PP059587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2023 22:48
 Operator : YP\AJ
 Sample : 04169-01
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PL-01-082423

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/28/2023
 Supervised By :Ankita Jodhani 08/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 00:05:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 24 01:16:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.639	26246539	43857792	22.206	22.701
2) SA Decachlor...	10.195	8.664	22327791	33701023	28.556	24.418
Target Compounds						
26) L6 AR-1254-1	6.441	5.533	1474995	2355959	30.407	22.956
27) L6 AR-1254-2	6.666	5.681	1684881	1515952	23.361m	17.290m#
28) L6 AR-1254-3	7.035	6.085	2048410	3887423	28.883m	27.723m
29) L6 AR-1254-4	7.319	6.314	1637979	2229754	37.027m	28.119m
30) L6 AR-1254-5	7.741	6.734	3050730	9225570	66.596	76.656m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082523\
Data File : PP059587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2023 22:48
Operator : YP\AJ
Sample : 04169-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PL-01-082423

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/28/2023

Supervised By :Ankita Jodhani 08/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 00:05:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 24 01:16:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

