

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082523\
 Data File : PP059579.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2023 20:37
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 22:32:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 24 01:16:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.412	3.640	51409014	90557314	43.495	46.874
2) SA Decachlor...	10.197	8.665	39795748	64138980	50.896	46.472
Target Compounds						
16) L4 AR-1242-1	5.588	4.729	15391661	23980981	457.648	467.174
17) L4 AR-1242-2	5.610	4.748	21902679	34899739	449.286	459.850
18) L4 AR-1242-3	5.673	4.923	14027159	18267716	452.685	472.092
19) L4 AR-1242-4	5.772	5.008	11351945	18730686	455.455	459.139
20) L4 AR-1242-5	6.512	5.534	12194990	21984827	470.263	468.441



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082523\
Data File : PP059579.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2023 20:37
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 22:32:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 24 01:16:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.412 min
Delta R.T.: 0.001 min
Response: 51409014
Conc: 43.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: -0.001 min
Response: 90557314
Conc: 46.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.197 min
Delta R.T.: 0.002 min
Response: 39795748
Conc: 50.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.665 min
Delta R.T.: -0.005 min
Response: 64138980
Conc: 46.47 ng/ml



#16 AR-1242-1


R.T.: 5.588 min
Delta R.T.: 0.000 min
Response: 15391661
Conc: 457.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.729 min
Delta R.T.: -0.003 min
Response: 23980981
Conc: 467.17 ng/ml




#17 AR-1242-2

R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 21902679
Conc: 449.29 ng/ml



#17 AR-1242-2


R.T.: 4.748 min
Delta R.T.: -0.004 min
Response: 34899739
Conc: 459.85 ng/ml



#18 AR-1242-3


R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 14027159
Conc: 452.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.923 min
Delta R.T.: -0.004 min
Response: 18267716
Conc: 472.09 ng/ml




#19 AR-1242-4

R.T.: 5.772 min
Delta R.T.: 0.000 min
Response: 11351945
Conc: 455.46 ng/ml



#19 AR-1242-4


R.T.: 5.008 min
Delta R.T.: -0.004 min
Response: 18730686
Conc: 459.14 ng/ml



#20 AR-1242-5

R.T.: 6.512 min
Delta R.T.: 0.000 min
Response: 12194990
Conc: 470.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.534 min
Delta R.T.: -0.003 min
Response: 21984827
Conc: 468.44 ng/ml