

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082621\
 Data File : PP038935.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Aug 2021 9:29
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 13:13:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 26 09:31:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.884	3.861	2039718	1309587	53.436	54.341
2)	SA Decachlor...	10.867	9.120	1323888	1243362	54.985	54.988
Target Compounds							
3)	L1 AR-1016-1	6.203	5.105	634851	458167	550.653	529.821
4)	L1 AR-1016-2	6.227	5.124	908625	659590	543.850	545.656
5)	L1 AR-1016-3	6.293	5.314	564238	352537	560.009	546.092
6)	L1 AR-1016-4	6.400	5.369	467889	270296	565.482	554.266
7)	L1 AR-1016-5	6.715	5.595	422058	342114	590.386	542.699
31)	L7 AR-1260-1	7.890	6.688	526485	608933	552.337	555.574
32)	L7 AR-1260-2	8.156	6.887	625470	719471	519.295	540.770
33)	L7 AR-1260-3	8.521	7.038	463004	688840	525.109	498.474
34)	L7 AR-1260-4	8.752	7.520	551367	584938	544.810	556.364
35)	L7 AR-1260-5	9.075	7.769	1137712	1404273	524.999	540.986



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082621\
Data File : PP038935.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Aug 2021 9:29
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 13:13:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 26 09:31:39 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.884 min
Delta R.T.: 0.007 min
Response: 2039718
Conc: 53.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.861 min
Delta R.T.: 0.000 min
Response: 1309587
Conc: 54.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.867 min
Delta R.T.: 0.009 min
Response: 1323888
Conc: 54.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.120 min
Delta R.T.: 0.000 min
Response: 1243362
Conc: 54.99 ng/ml



#3 AR-1016-1


R.T.: 6.203 min
Delta R.T.: 0.007 min
Response: 634851
Conc: 550.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.105 min
Delta R.T.: 0.000 min
Response: 458167
Conc: 529.82 ng/ml




#4 AR-1016-2

R.T.: 6.227 min
Delta R.T.: 0.008 min
Response: 908625
Conc: 543.85 ng/ml



#4 AR-1016-2


R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 659590
Conc: 545.66 ng/ml



#5 AR-1016-3


R.T.: 6.293 min
Delta R.T.: 0.007 min
Response: 564238
Conc: 560.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.314 min
Delta R.T.: 0.000 min
Response: 352537
Conc: 546.09 ng/ml




#6 AR-1016-4

R.T.: 6.400 min
Delta R.T.: 0.007 min
Response: 467889
Conc: 565.48 ng/ml



#6 AR-1016-4


R.T.: 5.369 min
Delta R.T.: 0.000 min
Response: 270296
Conc: 554.27 ng/ml



#7 AR-1016-5


R.T.: 6.715 min
Delta R.T.: 0.008 min
Response: 422058
Conc: 590.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 342114
Conc: 542.70 ng/ml




#31 AR-1260-1

R.T.: 7.890 min
Delta R.T.: 0.007 min
Response: 526485
Conc: 552.34 ng/ml



#31 AR-1260-1


R.T.: 6.688 min
Delta R.T.: 0.000 min
Response: 608933
Conc: 555.57 ng/ml



#32 AR-1260-2


R.T.: 8.156 min
Delta R.T.: 0.006 min
Response: 625470
Conc: 519.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.887 min
Delta R.T.: 0.000 min
Response: 719471
Conc: 540.77 ng/ml




#33 AR-1260-3

R.T.: 8.521 min
Delta R.T.: 0.007 min
Response: 463004
Conc: 525.11 ng/ml



#33 AR-1260-3


R.T.: 7.038 min
Delta R.T.: 0.000 min
Response: 688840
Conc: 498.47 ng/ml



#34 AR-1260-4


R.T.: 8.752 min
Delta R.T.: 0.006 min
Response: 551367
Conc: 544.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 584938
Conc: 556.36 ng/ml



#35 AR-1260-5

R.T.: 9.075 min
Delta R.T.: 0.007 min
Response: 1137712
Conc: 525.00 ng/ml



#35 AR-1260-5

R.T.: 7.769 min
Delta R.T.: 0.000 min
Response: 1404273
Conc: 540.99 ng/ml