

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082722\
 Data File : PP050917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Aug 2022 04:30
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/29/2022
 Supervised By :Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 00:50:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.461	3.718	98724561	91266954	52.977	51.793
2) SA Decachlor...	10.258	8.806	96914459	58082373	53.723	54.369
Target Compounds						
3) L1 AR-1016-1	5.632	4.819	32809107	25986359	543.185	532.314m
4) L1 AR-1016-2	5.654	4.839	47605594	37884891	541.392	539.192m
5) L1 AR-1016-3	5.717	5.018	30609771	20391528	541.974	520.501m
6) L1 AR-1016-4	5.815	5.059	24992620	16810412	558.899	508.704m
7) L1 AR-1016-5	6.111	5.276	26936678	20964241	537.680m	490.724m
31) L7 AR-1260-1	7.239	6.319	50875429	40067179	537.292m	530.052m
32) L7 AR-1260-2	7.496	6.506	56925081	48745142	562.658m	580.437m
33) L7 AR-1260-3	7.856	6.663	43231744	44761768	527.318m	538.041m
34) L7 AR-1260-4	8.083	7.139	49254107	35243958	544.325	552.431m
35) L7 AR-1260-5	8.407	7.379	87268607	72915559	569.685m	582.658m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082722\
Data File : PP050917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Aug 2022 04:30
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/29/2022
Supervised By :Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 00:50:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

