

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082722\
 Data File : PP050899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Aug 2022 21:49
 Operator : YP\AJ
 Sample : N4330-05
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 OG-WATER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 00:43:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.465	3.723	43502809	40066459	23.344	22.737
2) SA Decachlor...	10.262	8.810	43764119	23193688	24.260	21.711

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082722\
Data File : PP050899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Aug 2022 21:49
Operator : YP\AJ
Sample : N4330-05
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OG-WATER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 00:43:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.465 min
Delta R.T.: 0.003 min
Response: 43502809
Conc: 23.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
OG-WATER




#1 Tetrachloro-m-xylene

R.T.: 3.723 min
Delta R.T.: 0.003 min
Response: 40066459
Conc: 22.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.262 min
Delta R.T.: 0.002 min
Response: 43764119
Conc: 24.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.810 min
Delta R.T.: -0.002 min
Response: 23193688
Conc: 21.71 ng/ml