

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082724\
 Data File : PP066997.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2024 06:53
 Operator : YP\AJ
 Sample : P3761-07
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PIPE-7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 07:18:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 09:05:47 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.623	3.911	34469772	28279771	20.479	19.729
2)	SA Decachlor...	10.482	9.048	19024625	17130402	22.688	22.872
Target Compounds							
3)	L1 AR-1016-1	5.801	5.015	4687516	11593730	99.071	265.586 #
4)	L1 AR-1016-2	5.811	5.042	8629831	6688959	116.636	103.948
5)	L1 AR-1016-3	5.884	5.236	1349452	8346937	28.762	243.071 #
6)	L1 AR-1016-4	5.982	5.256	2219754	8903110	61.394	293.925 #
7)	L1 AR-1016-5	6.279	0.000	1311068	0	38.271	N.D. #
8)	L2 AR-1221-1	4.823	4.128	113260	182147	6.243	11.712 #
9)	L2 AR-1221-2	4.929	4.166f	483830	109056	35.939	9.321 #
10)	L2 AR-1221-3	4.999	4.276	181983	5340448	4.503	140.262 #
11)	L3 AR-1232-1	4.999	4.276	181983	5340448	5.459	169.682 #
12)	L3 AR-1232-2	5.528	5.042	5729244	6688959	324.540	226.831 #
13)	L3 AR-1232-3	5.811	5.236	8629831	8346937	276.385	587.722 #
14)	L3 AR-1232-4	5.982	5.313	2219754	10551462	145.081	750.675 #
15)	L3 AR-1232-5	6.067	0.000	644147	0	51.192	N.D. #
16)	L4 AR-1242-1	5.801	5.015	4687516	11593730	136.276	373.205 #
17)	L4 AR-1242-2	5.811	5.042	8629831	6688959	162.467	149.864
18)	L4 AR-1242-3	5.884	5.236	1349452	8346937	39.351	346.018 #
19)	L4 AR-1242-4	5.982	5.313	2219754	10551462	85.541	394.298 #
20)	L4 AR-1242-5	6.728	5.834	3802403	2890721	163.644	103.572 #
21)	L5 AR-1248-1	5.801	5.015	4687516	11593730	194.179	526.766 #
22)	L5 AR-1248-2	6.067	5.256	644147	8903110	16.539	253.414 #
23)	L5 AR-1248-3	6.279	5.313	1311068	10551462	32.187	289.971 #
24)	L5 AR-1248-4	6.679	0.000	1944561	0	50.189	N.D. #
25)	L5 AR-1248-5	6.728	5.870	3802403	522501	96.561	16.356 #
26)	L6 AR-1254-1	6.654	5.834	2863753	2890721	57.554	48.712
27)	L6 AR-1254-2	6.871	5.983	8250024	5053239	116.796	96.438
28)	L6 AR-1254-3	7.239	6.394	12725918	13808672	186.776	181.009
29)	L6 AR-1254-4	7.523	6.623	12652708	11137912	302.168	294.849
30)	L6 AR-1254-5	7.942	7.046	14557391	15826501	277.043	251.465
31)	L7 AR-1260-1	7.402	6.525	6791625	8439875	112.216	133.230
32)	L7 AR-1260-2	7.657	6.713	8296318	7796843	129.060	115.312
33)	L7 AR-1260-3	8.003	6.870	1717359	8273619	38.472	124.221 #
34)	L7 AR-1260-4	8.237	7.346	3161061	841416	63.092	19.198 #
35)	L7 AR-1260-5	8.581	7.585	1478180	1184778	17.687	13.284
36)	L8 AR-1262-1	8.237	7.046f	3161061	15826501	48.337	219.066 #
37)	L8 AR-1262-2	8.581	7.346	1478180	841416	14.291	13.674
38)	L8 AR-1262-3	8.923	7.867	908005	1243585	12.714	29.564 #
39)	L8 AR-1262-4	9.007	7.937	58980	1177206	0.999	15.176 #
40)	L8 AR-1262-5	9.697	8.459	37537	57150	1.364	2.164 #
41)	L9 AR-1268-1	8.923	7.867	908005	1243585	6.948	9.942 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082724\
Data File : PP066997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2024 06:53
Operator : YP\AJ
Sample : P3761-07
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PIPE-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 07:18:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 09:05:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.007	7.937	58980	1177206	0.504	10.533 #
43) L9	AR-1268-3	9.242	0.000	304939	0	2.930	N.D. #
44) L9	AR-1268-4	9.697	8.459	37537	57150	1.254	2.013 #
45) L9	AR-1268-5	0.000	8.777	0	414808	N.D.	1.691 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082724\
Data File : PP066997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2024 06:53
Operator : YP\AJ
Sample : P3761-07
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PIPE-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 07:18:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 09:05:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.623 min
Delta R.T.: -0.004 min
Response: 34469772
Conc: 20.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#1 Tetrachloro-m-xylene

R.T.: 3.911 min
Delta R.T.: -0.003 min
Response: 28279771
Conc: 19.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.482 min
Delta R.T.: -0.009 min
Response: 19024625
Conc: 22.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.048 min
Delta R.T.: -0.005 min
Response: 17130402
Conc: 22.87 ng/ml



#3 AR-1016-1


R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 4687516
Conc: 99.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#3 AR-1016-1

R.T.: 5.015 min
Delta R.T.: -0.008 min
Response: 11593730
Conc: 265.59 ng/ml




#4 AR-1016-2

R.T.: 5.811 min
Delta R.T.: -0.012 min
Response: 8629831
Conc: 116.64 ng/ml



#4 AR-1016-2


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 6688959
Conc: 103.95 ng/ml



#5 AR-1016-3


R.T.: 5.884 min
Delta R.T.: -0.002 min
Response: 1349452
Conc: 28.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#5 AR-1016-3

R.T.: 5.236 min
Delta R.T.: 0.012 min
Response: 8346937
Conc: 243.07 ng/ml




#6 AR-1016-4

R.T.: 5.982 min
Delta R.T.: -0.004 min
Response: 2219754
Conc: 61.39 ng/ml



#6 AR-1016-4


R.T.: 5.256 min
Delta R.T.: -0.007 min
Response: 8903110
Conc: 293.93 ng/ml



#7 AR-1016-5


R.T.: 6.279 min
Delta R.T.: 0.000 min
Response: 1311068
Conc: 38.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.483 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.823 min
Delta R.T.: -0.011 min
Response: 113260
Conc: 6.24 ng/ml



#8 AR-1221-1


R.T.: 4.128 min
Delta R.T.: -0.002 min
Response: 182147
Conc: 11.71 ng/ml



#9 AR-1221-2


R.T.: 4.929 min
Delta R.T.: 0.008 min
Response: 483830
Conc: 35.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#9 AR-1221-2

R.T.: 4.166 min
Delta R.T.: -0.053 min
Response: 109056
Conc: 9.32 ng/ml




#10 AR-1221-3

R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 181983
Conc: 4.50 ng/ml



#10 AR-1221-3


R.T.: 4.276 min
Delta R.T.: -0.021 min
Response: 5340448
Conc: 140.26 ng/ml



#11 AR-1232-1


R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 181983
Conc: 5.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#11 AR-1232-1

R.T.: 4.276 min
Delta R.T.: -0.023 min
Response: 5340448
Conc: 169.68 ng/ml




#12 AR-1232-2

R.T.: 5.528 min
Delta R.T.: -0.005 min
Response: 5729244
Conc: 324.54 ng/ml



#12 AR-1232-2


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 6688959
Conc: 226.83 ng/ml



#13 AR-1232-3


R.T.: 5.811 min
Delta R.T.: -0.012 min
Response: 8629831
Conc: 276.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#13 AR-1232-3

R.T.: 5.236 min
Delta R.T.: 0.012 min
Response: 8346937
Conc: 587.72 ng/ml




#14 AR-1232-4

R.T.: 5.982 min
Delta R.T.: -0.004 min
Response: 2219754
Conc: 145.08 ng/ml



#14 AR-1232-4


R.T.: 5.313 min
Delta R.T.: 0.005 min
Response: 10551462
Conc: 750.67 ng/ml



#15 AR-1232-5


R.T.: 6.067 min
Delta R.T.: -0.007 min
Response: 644147
Conc: 51.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.483 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.801 min
Delta R.T.: -0.001 min
Response: 4687516
Conc: 136.28 ng/ml



#16 AR-1242-1


R.T.: 5.015 min
Delta R.T.: -0.010 min
Response: 11593730
Conc: 373.21 ng/ml



#17 AR-1242-2


R.T.: 5.811 min
Delta R.T.: -0.013 min
Response: 8629831
Conc: 162.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#17 AR-1242-2

R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 6688959
Conc: 149.86 ng/ml




#18 AR-1242-3

R.T.: 5.884 min
Delta R.T.: -0.003 min
Response: 1349452
Conc: 39.35 ng/ml



#18 AR-1242-3


R.T.: 5.236 min
Delta R.T.: 0.012 min
Response: 8346937
Conc: 346.02 ng/ml



#19 AR-1242-4


R.T.: 5.982 min
Delta R.T.: -0.004 min
Response: 2219754
Conc: 85.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#19 AR-1242-4

R.T.: 5.313 min
Delta R.T.: 0.005 min
Response: 10551462
Conc: 394.30 ng/ml




#20 AR-1242-5

R.T.: 6.728 min
Delta R.T.: 0.005 min
Response: 3802403
Conc: 163.64 ng/ml



#20 AR-1242-5


R.T.: 5.834 min
Delta R.T.: -0.004 min
Response: 2890721
Conc: 103.57 ng/ml



#21 AR-1248-1


R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 4687516
Conc: 194.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#21 AR-1248-1

R.T.: 5.015 min
Delta R.T.: -0.009 min
Response: 11593730
Conc: 526.77 ng/ml




#22 AR-1248-2

R.T.: 6.067 min
Delta R.T.: -0.008 min
Response: 644147
Conc: 16.54 ng/ml



#22 AR-1248-2


R.T.: 5.256 min
Delta R.T.: -0.007 min
Response: 8903110
Conc: 253.41 ng/ml



#23 AR-1248-3


R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 1311068
Conc: 32.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#23 AR-1248-3

R.T.: 5.313 min
Delta R.T.: 0.006 min
Response: 10551462
Conc: 289.97 ng/ml




#24 AR-1248-4

R.T.: 6.679 min
Delta R.T.: -0.006 min
Response: 1944561
Conc: 50.19 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.482 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.728 min
Delta R.T.: 0.004 min
Response: 3802403
Conc: 96.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#25 AR-1248-5

R.T.: 5.870 min
Delta R.T.: -0.011 min
Response: 522501
Conc: 16.36 ng/ml




#26 AR-1254-1

R.T.: 6.654 min
Delta R.T.: -0.002 min
Response: 2863753
Conc: 57.55 ng/ml



#26 AR-1254-1


R.T.: 5.834 min
Delta R.T.: -0.003 min
Response: 2890721
Conc: 48.71 ng/ml



#27 AR-1254-2


R.T.: 6.871 min
Delta R.T.: -0.004 min
Response: 8250024
Conc: 116.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#27 AR-1254-2

R.T.: 5.983 min
Delta R.T.: -0.003 min
Response: 5053239
Conc: 96.44 ng/ml




#28 AR-1254-3

R.T.: 7.239 min
Delta R.T.: -0.003 min
Response: 12725918
Conc: 186.78 ng/ml



#28 AR-1254-3


R.T.: 6.394 min
Delta R.T.: -0.003 min
Response: 13808672
Conc: 181.01 ng/ml



#29 AR-1254-4


R.T.: 7.523 min
Delta R.T.: -0.003 min
Response: 12652708
Conc: 302.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#29 AR-1254-4

R.T.: 6.623 min
Delta R.T.: -0.002 min
Response: 11137912
Conc: 294.85 ng/ml




#30 AR-1254-5

R.T.: 7.942 min
Delta R.T.: -0.002 min
Response: 14557391
Conc: 277.04 ng/ml



#30 AR-1254-5


R.T.: 7.046 min
Delta R.T.: -0.003 min
Response: 15826501
Conc: 251.46 ng/ml



#31 AR-1260-1


R.T.: 7.402 min
Delta R.T.: -0.003 min
Response: 6791625
Conc: 112.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#31 AR-1260-1

R.T.: 6.525 min
Delta R.T.: -0.004 min
Response: 8439875
Conc: 133.23 ng/ml




#32 AR-1260-2

R.T.: 7.657 min
Delta R.T.: -0.004 min
Response: 8296318
Conc: 129.06 ng/ml



#32 AR-1260-2


R.T.: 6.713 min
Delta R.T.: -0.004 min
Response: 7796843
Conc: 115.31 ng/ml



#33 AR-1260-3


R.T.: 8.003 min
Delta R.T.: -0.018 min
Response: 1717359
Conc: 38.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#33 AR-1260-3

R.T.: 6.870 min
Delta R.T.: -0.005 min
Response: 8273619
Conc: 124.22 ng/ml




#34 AR-1260-4

R.T.: 8.237 min
Delta R.T.: -0.017 min
Response: 3161061
Conc: 63.09 ng/ml



#34 AR-1260-4


R.T.: 7.346 min
Delta R.T.: -0.005 min
Response: 841416
Conc: 19.20 ng/ml



#35 AR-1260-5


R.T.: 8.581 min
Delta R.T.: -0.006 min
Response: 1478180
Conc: 17.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#35 AR-1260-5

R.T.: 7.585 min
Delta R.T.: -0.006 min
Response: 1184778
Conc: 13.28 ng/ml




#36 AR-1262-1

R.T.: 8.237 min
Delta R.T.: -0.015 min
Response: 3161061
Conc: 48.34 ng/ml



#36 AR-1262-1


R.T.: 7.046 min
Delta R.T.: -0.039 min
Response: 15826501
Conc: 219.07 ng/ml



#37 AR-1262-2


R.T.: 8.581 min
Delta R.T.: -0.001 min
Response: 1478180
Conc: 14.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#37 AR-1262-2

R.T.: 7.346 min
Delta R.T.: -0.002 min
Response: 841416
Conc: 13.67 ng/ml




#38 AR-1262-3

R.T.: 8.923 min
Delta R.T.: 0.013 min
Response: 908005
Conc: 12.71 ng/ml



#38 AR-1262-3


R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 1243585
Conc: 29.56 ng/ml



#39 AR-1262-4


R.T.: 9.007 min
Delta R.T.: 0.003 min
Response: 58980
Conc: 1.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#39 AR-1262-4

R.T.: 7.937 min
Delta R.T.: -0.003 min
Response: 1177206
Conc: 15.18 ng/ml




#40 AR-1262-5

R.T.: 9.697 min
Delta R.T.: 0.006 min
Response: 37537
Conc: 1.36 ng/ml



#40 AR-1262-5


R.T.: 8.459 min
Delta R.T.: 0.004 min
Response: 57150
Conc: 2.16 ng/ml



#41 AR-1268-1


R.T.: 8.923 min
Delta R.T.: 0.015 min
Response: 908005
Conc: 6.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#41 AR-1268-1

R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 1243585
Conc: 9.94 ng/ml




#42 AR-1268-2

R.T.: 9.007 min
Delta R.T.: 0.000 min
Response: 58980
Conc: 0.50 ng/ml



#42 AR-1268-2


R.T.: 7.937 min
Delta R.T.: -0.004 min
Response: 1177206
Conc: 10.53 ng/ml



#43 AR-1268-3


R.T.: 9.242 min
Delta R.T.: -0.004 min
Response: 304939
Conc: 2.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
PIPE-7




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.151 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.697 min
Delta R.T.: 0.006 min
Response: 37537
Conc: 1.25 ng/ml



#44 AR-1268-4


R.T.: 8.459 min
Delta R.T.: 0.003 min
Response: 57150
Conc: 2.01 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.127 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
PIPE-7



#45 AR-1268-5

R.T.: 8.777 min
Delta R.T.: 0.005 min
Response: 414808
Conc: 1.69 ng/ml