

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP082918.M
 Title : 8080.M
 Last Update : Thu Aug 30 07:11:19 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP027230.D 1000 =PP027231.D 1500 =PP027232.D
 500 =PP027229.D 200 =PP027228.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.231	6.067	5.970	6.164	6.194	6.125	E5	1.73
2) T	3,5-DICHLOROBENZOI...	4.770	4.616	4.464	4.768	4.751	4.674	E5	2.86
3) T	4-Nitrophenol	4.334	4.283	4.234	4.286	4.026	4.233	E5	2.85
4) S	2,4-DCAA	2.542	2.478	2.435	2.512	2.440	2.482	E5	1.85
5) T	DICAMBA	1.566	1.517	1.467	1.557	1.599	1.541	E6	3.29
6) T	MCPP	4.339	4.416	4.806	3.596	2.431	3.918	E5	23.98
7) T	MCPA	7.463	7.647	8.112	5.817	3.753	6.558	E5	27.32
8) T	DICHLORPROP	3.323	3.231	3.113	3.307	3.275	3.250	E5	2.59
9) T	2,4-D	3.882	3.824	3.761	3.791	3.467	3.745	E5	4.32
10) T	Pentachlorophenol	8.203	7.754	7.163	8.412	8.723	8.051	E6	7.56
11) T	2,4,5-TP (SILVEX)	2.967	2.847	2.717	2.994	3.064	2.918	E6	4.69
12) T	2,4,5-T	2.519	2.458	2.383	2.500	2.392	2.451	E6	2.51
13) T	2,4-DB	2.002	2.007	2.061	2.013	1.605	1.938	E5	9.68
14) T	DINOSEB	2.091	2.029	1.967	2.095	2.094	2.055	E6	2.76
15) T	Picloram	3.386	3.489	3.547	3.186	2.640	3.250	E6	11.31
16) T	DCPA	4.304	4.120	3.922	4.354	4.427	4.225	E6	4.82

Signal #2 Calibration Files

750 =PP027230.D 1000 =PP027231.D 1500 =PP027232.D
 500 =PP027229.D 200 =PP027228.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.103	5.957	5.805	6.054	6.028	5.989	E5	1.93
2) T	3,5-DICHLOROBENZOI...	5.581	5.418	5.246	5.630	5.721	5.519	E5	3.41
3) T	4-Nitrophenol	2.396	2.354	2.333	2.366	2.102	2.310	E5	5.14
4) S	2,4-DCAA	4.572	4.439	4.314	4.610	4.676	4.522	E5	3.21
5) T	DICAMBA	1.748	1.683	1.607	1.771	1.829	1.727	E6	4.94
6) T	MCPP	8.924	8.857	9.097	7.942	6.221	8.208	E5	14.60
7) T	MCPA	1.388	1.400	1.414	1.318	1.012	1.307	E6	12.91
8) T	DICHLORPROP	4.297	4.168	4.010	4.379	4.474	4.265	E5	4.25
9) T	2,4-D	5.560	5.438	5.272	5.631	5.359	5.452	E5	2.67
10) T	Pentachlorophenol	6.923	6.472	5.860	7.347	7.880	6.896	E6	11.29
11) T	2,4,5-TP (SILVEX)	2.629	2.520	2.361	2.719	2.785	2.603	E6	6.45
12) T	2,4,5-T	2.113	2.066	1.991	2.138	1.980	2.057	E6	3.44
13) T	2,4-DB	2.693	2.701	2.686	2.643	2.351	2.615	E5	5.70
14) T	DINOSEB	1.856	1.788	1.708	1.911	1.908	1.834	E6	4.71
15) T	Picloram	2.896	2.928	2.923	2.816	2.277	2.768	E6	10.05
16) T	DCPA	3.349	3.162	2.931	3.507	3.681	3.326	E6	8.80

(#) = Out of Range