

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082922\
 Data File : PP050999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2022 01:10
 Operator : YP\AJ
 Sample : N4398-01
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ROCK-FINES-QA/QC-COMP

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/30/2022
 Supervised By :Ankita Jodhani 08/30/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 03:38:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.461	3.716	40362655	69295201	21.659m	39.324 #
2) SA Decachlor...	10.257	8.799	48681990	28669492	26.986	26.837

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082922\
Data File : PP050999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2022 01:10
Operator : YP\AJ
Sample : N4398-01
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ROCK-FINES-QA/QC-COMP


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/30/2022
Supervised By :Ankita Jodhani 08/30/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 03:38:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.461 min
Delta R.T.: -0.001 min
Response: 40362655
Conc: 21.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ROCK-FINES-QA/QC-COMP


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/30/2022
Supervised By :Ankita Jodhani 08/30/2022




#1 Tetrachloro-m-xylene

R.T.: 3.716 min
Delta R.T.: -0.005 min
Response: 69295201
Conc: 39.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.257 min
Delta R.T.: -0.003 min
Response: 48681990
Conc: 26.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.799 min
Delta R.T.: -0.012 min
Response: 28669492
Conc: 26.84 ng/ml