

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082922\
 Data File : PP051014.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2022 06:51
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 07:14:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.462	3.718	105.2E6	99533294	56.427	56.484
2) SA Decachlor...	10.259	8.799	109.9E6	65370375	60.933	61.191
Target Compounds						
26) L6 AR-1254-1	6.490	5.629	49529742	46817736	621.466	581.675
27) L6 AR-1254-2	6.709	5.776	63275325	37532467	535.856	537.974
28) L6 AR-1254-3	7.076	6.185	69278288	61182729	605.462	571.109
29) L6 AR-1254-4	7.361	6.413	47761761	34565649	693.554	691.885
30) L6 AR-1254-5	7.782	6.836	61885813	52870448	741.306	622.695



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082922\
Data File : PP051014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2022 06:51
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 07:14:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.462 min
Delta R.T.: 0.000 min
Response: 105153534
Conc: 56.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.003 min
Response: 99533294
Conc: 56.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.259 min
Delta R.T.: -0.002 min
Response: 109921072
Conc: 60.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: -0.012 min
Response: 65370375
Conc: 61.19 ng/ml



#26 AR-1254-1


R.T.: 6.490 min
Delta R.T.: -0.003 min
Response: 49529742
Conc: 621.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.629 min
Delta R.T.: -0.009 min
Response: 46817736
Conc: 581.67 ng/ml




#27 AR-1254-2

R.T.: 6.709 min
Delta R.T.: -0.003 min
Response: 63275325
Conc: 535.86 ng/ml



#27 AR-1254-2


R.T.: 5.776 min
Delta R.T.: -0.009 min
Response: 37532467
Conc: 537.97 ng/ml



#28 AR-1254-3


R.T.: 7.076 min
Delta R.T.: -0.002 min
Response: 69278288
Conc: 605.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.185 min
Delta R.T.: -0.009 min
Response: 61182729
Conc: 571.11 ng/ml




#29 AR-1254-4

R.T.: 7.361 min
Delta R.T.: -0.003 min
Response: 47761761
Conc: 693.55 ng/ml



#29 AR-1254-4


R.T.: 6.413 min
Delta R.T.: -0.009 min
Response: 34565649
Conc: 691.88 ng/ml



#30 AR-1254-5

R.T.: 7.782 min
Delta R.T.: -0.002 min
Response: 61885813
Conc: 741.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.836 min
Delta R.T.: -0.009 min
Response: 52870448
Conc: 622.70 ng/ml