

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083022\
 Data File : PP051018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2022 09:33
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 16:40:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.465	3.717	88738767	85876820	47.618	48.734
2) SA Decachlor...	10.265	8.800	93625443	56873179	51.899	53.237
Target Compounds						
16) L4 AR-1242-1	5.636	4.817	25073958	20522059	507.369	511.587
17) L4 AR-1242-2	5.658	4.837	36108492	28914878	505.673	507.132
18) L4 AR-1242-3	5.721	5.016	23055119	16464681	503.522	516.965
19) L4 AR-1242-4	5.820	5.099	18317965	16650144	507.918	506.425
20) L4 AR-1242-5	6.558	5.627	19062058	18586370	477.062	537.479



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083022\
Data File : PP051018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2022 09:33
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 16:40:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.465 min
Delta R.T.: 0.003 min
Response: 88738767
Conc: 47.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: -0.003 min
Response: 85876820
Conc: 48.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.265 min
Delta R.T.: 0.005 min
Response: 93625443
Conc: 51.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.800 min
Delta R.T.: -0.011 min
Response: 56873179
Conc: 53.24 ng/ml



#16 AR-1242-1


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 25073958
Conc: 507.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.817 min
Delta R.T.: -0.008 min
Response: 20522059
Conc: 511.59 ng/ml




#17 AR-1242-2

R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 36108492
Conc: 505.67 ng/ml



#17 AR-1242-2


R.T.: 4.837 min
Delta R.T.: -0.008 min
Response: 28914878
Conc: 507.13 ng/ml



#18 AR-1242-3


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 23055119
Conc: 503.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.016 min
Delta R.T.: -0.009 min
Response: 16464681
Conc: 516.96 ng/ml




#19 AR-1242-4

R.T.: 5.820 min
Delta R.T.: 0.000 min
Response: 18317965
Conc: 507.92 ng/ml



#19 AR-1242-4


R.T.: 5.099 min
Delta R.T.: -0.009 min
Response: 16650144
Conc: 506.43 ng/ml



#20 AR-1242-5

R.T.: 6.558 min
Delta R.T.: 0.000 min
Response: 19062058
Conc: 477.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.627 min
Delta R.T.: -0.010 min
Response: 18586370
Conc: 537.48 ng/ml