

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083022\
 Data File : PP051059.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 00:10
 Operator : YP\AJ
 Sample : N4432-06 10X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 KMA9320-2-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 02:23:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.461	3.717	2558098	2875035	1.373	1.632
2) SA Decachlor...	10.260	8.799	3934852	2491446	2.181	2.332

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083022\
Data File : PP051059.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 00:10
Operator : YP\AJ
Sample : N4432-06 10X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
KMA9320-2-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 02:23:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.461 min
Delta R.T.: 0.000 min
Response: 2558098
Conc: 1.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
KMA9320-2-2




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: -0.003 min
Response: 2875035
Conc: 1.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: 0.000 min
Response: 3934852
Conc: 2.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.799 min
Delta R.T.: -0.013 min
Response: 2491446
Conc: 2.33 ng/ml