

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083122\
 Data File : PP051108.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 23:39
 Operator : YP\AJ
 Sample : N4448-16
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 40075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 03:57:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 26 17:33:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.462	3.717	34497998	36859914	18.512	20.918
2) SA Decachlor...	10.261	8.797	43755422	27945873	24.255	26.159

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP083122\
Data File : PP051108.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 23:39
Operator : YP\AJ
Sample : N4448-16
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
40075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 03:57:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 26 17:33:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.462 min
Delta R.T.: 0.000 min
Response: 34497998
Conc: 18.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
40075




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: -0.003 min
Response: 36859914
Conc: 20.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: 0.001 min
Response: 43755422
Conc: 24.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: -0.014 min
Response: 27945873
Conc: 26.16 ng/ml