

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090221\
 Data File : PP039163.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:33
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 02:51:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	38.311	39.077 E3	-2.0	101	0.00
2 SA Decachlorobiphenyl	32.916	37.490 E3	-13.9	111	0.00
16 L4 AR-1242-1	979.387	1014.750	-3.6	102	0.00
17 L4 AR-1242-2	1.545	1.584 E3	-2.5	102	0.00
18 L4 AR-1242-3	0.986	1.016 E3	-3.0	100	0.00
19 L4 AR-1242-4	769.740	811.100	-5.4	100	0.00
20 L4 AR-1242-5	815.987	888.632	-8.9	107	0.00

Signal #2

1 SA Tetrachloro-m-xylene	23.986	24.101 E3	-0.5	97	0.00
2 SA Decachlorobiphenyl	21.885	24.361 E3	-11.3	110	0.00
16 L4 AR-1242-1	547.601	566.979	-3.5	100	0.00
17 L4 AR-1242-2	869.051	918.769	-5.7	102	0.00
18 L4 AR-1242-3	446.372	472.278	-5.8	100	0.00
19 L4 AR-1242-4	439.390	465.815	-6.0	99	0.00
20 L4 AR-1242-5	546.277	593.304	-8.6	110	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	1.275	0.000 E3	100.0#	0#	-6.19#
4 L1 AR-1016-2	2.026	0.000 E3	100.0#	0#	-6.21#
5 L1 AR-1016-3	1.294	0.000 E3	100.0#	0#	-6.28#
6 L1 AR-1016-4	1.009	0.000 E3	100.0#	0#	-6.39#
7 L1 AR-1016-5	965.396	0.000	100.0#	0#	-6.70#
8 L2 AR-1221-1	482.533	0.000	100.0#	0#	-5.12#
9 L2 AR-1221-2	371.361	0.000	100.0#	0#	-5.22#
10 L2 AR-1221-3	1.115	0.000 E3	100.0#	0#	-5.31#
11 L3 AR-1232-1	857.952	0.000	100.0#	0#	-5.31#
12 L3 AR-1232-2	493.651	0.000	100.0#	0#	-5.89#
13 L3 AR-1232-3	939.090	0.000	100.0#	0#	-6.21#
14 L3 AR-1232-4	469.743	0.000	100.0#	0#	-6.39#
15 L3 AR-1232-5	331.078	0.000	100.0#	0#	-6.48#
21 L5 AR-1248-1	766.627	0.000	100.0#	0#	-6.19#
22 L5 AR-1248-2	1.067	0.000 E3	100.0#	0#	-6.48#
23 L5 AR-1248-3	1.278	0.000 E3	100.0#	0#	-6.70#
24 L5 AR-1248-4	1.343	0.000 E3	100.0#	0#	-7.13#
25 L5 AR-1248-5	1.403	0.000 E3	100.0#	0#	-7.17#
26 L6 AR-1254-1	1.525	0.000 E3	100.0#	0#	-7.10#
27 L6 AR-1254-2	2.351	0.000 E3	100.0#	0#	-7.33#
28 L6 AR-1254-3	2.447	0.000 E3	100.0#	0#	-7.71#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090221\
 Data File : PP039163.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:33
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 02:51:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	1.660	0.000 E3	100.0#	0#	-8.01#
30 L6 AR-1254-5	1.994	0.000 E3	100.0#	0#	-8.43#
31 L7 AR-1260-1	1.718	0.000 E3	100.0#	0#	-7.88#
32 L7 AR-1260-2	1.986	0.000 E3	100.0#	0#	-8.14#
33 L7 AR-1260-3	1.350	0.000 E3	100.0#	0#	-8.51#
34 L7 AR-1260-4	1.612	0.000 E3	100.0#	0#	-8.74#
35 L7 AR-1260-5	2.770	0.000 E3	100.0#	0#	-9.06#
36 L8 AR-1262-1	2.228	0.000 E3	100.0#	0#	-8.50#
37 L8 AR-1262-2	3.703	0.000 E3	100.0#	0#	-9.06#
38 L8 AR-1262-3	1.852	0.000 E3	100.0#	0#	-9.36#
39 L8 AR-1262-4	1.288	0.000 E3	100.0#	0#	-9.45#
40 L8 AR-1262-5	1.356	0.000 E3	100.0#	0#	-10.11#
41 L9 AR-1268-1	4.905	0.000 E3	100.0#	0#	-9.36#
42 L9 AR-1268-2	4.657	0.000 E3	100.0#	0#	-9.46#
43 L9 AR-1268-3	4.054	0.000 E3	100.0#	0#	-9.68#
44 L9 AR-1268-4	1.644	0.000 E3	100.0#	0#	-10.11#
45 L9 AR-1268-5	12.678	0.000 E3	100.0#	0#	-10.51#

Signal #2

3 L1 AR-1016-1	721.942	0.000	100.0#	0#	-5.10#
4 L1 AR-1016-2	1.174	0.000 E3	100.0#	0#	-5.12#
5 L1 AR-1016-3	595.697	0.000	100.0#	0#	-5.31#
6 L1 AR-1016-4	482.936	0.000	100.0#	0#	-5.36#
7 L1 AR-1016-5	623.460	0.000	100.0#	0#	-5.59#
8 L2 AR-1221-1	286.414	0.000	100.0#	0#	-4.11#
9 L2 AR-1221-2	215.332	0.000	100.0#	0#	-4.21#
10 L2 AR-1221-3	709.607	0.000	100.0#	0#	-4.30#
11 L3 AR-1232-1	530.443	0.000	100.0#	0#	-4.30#
12 L3 AR-1232-2	532.780	0.000	100.0#	0#	-5.12#
13 L3 AR-1232-3	260.832	0.000	100.0#	0#	-5.31#
14 L3 AR-1232-4	226.243	0.000	100.0#	0#	-5.41#
15 L3 AR-1232-5	263.271	0.000	100.0#	0#	-5.59#
21 L5 AR-1248-1	421.135	0.000	100.0#	0#	-5.10#
22 L5 AR-1248-2	670.075	0.000	100.0#	0#	-5.36#
23 L5 AR-1248-3	667.757	0.000	100.0#	0#	-5.40#
24 L5 AR-1248-4	796.728	0.000	100.0#	0#	-5.59#
25 L5 AR-1248-5	696.542	0.000	100.0#	0#	-6.01#
26 L6 AR-1254-1	1.258	0.000 E3	100.0#	0#	-5.96#
27 L6 AR-1254-2	1.155	0.000 E3	100.0#	0#	-6.12#
28 L6 AR-1254-3	1.772	0.000 E3	100.0#	0#	-6.54#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090221\
 Data File : PP039163.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2021 2:33
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 02:51:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	913.833	0.000	100.0#	0#	-6.78#
30 L6 AR-1254-5	1.654	0.000 E3	100.0#	0#	-7.21#
31 L7 AR-1260-1	1.203	0.000 E3	100.0#	0#	-6.68#
32 L7 AR-1260-2	1.398	0.000 E3	100.0#	0#	-6.88#
33 L7 AR-1260-3	1.366	0.000 E3	100.0#	0#	-7.03#
34 L7 AR-1260-4	0.972	0.000 E3	100.0#	0#	-7.51#
35 L7 AR-1260-5	2.066	0.000 E3	100.0#	0#	-7.76#
36 L8 AR-1262-1	759.667	0.000	100.0#	0#	-7.21#
37 L8 AR-1262-2	2.575	0.000 E3	100.0#	0#	-7.76#
38 L8 AR-1262-3	1.139	0.000 E3	100.0#	0#	-8.04#
39 L8 AR-1262-4	2.048	0.000 E3	100.0#	0#	-8.11#
40 L8 AR-1262-5	886.266	0.000	100.0#	0#	-8.61#
41 L9 AR-1268-1	3.377	0.000 E3	100.0#	0#	-8.04#
42 L9 AR-1268-2	3.165	0.000 E3	100.0#	0#	-8.11#
43 L9 AR-1268-3	2.683	0.000 E3	100.0#	0#	-8.31#
44 L9 AR-1268-4	1.069	0.000 E3	100.0#	0#	-8.61#
45 L9 AR-1268-5	8.559	0.000 E3	100.0#	0#	-8.88#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0