

Data Path : P:\HPCHEM1\ECD P\Data\PP090315\
 Data File : PP011538.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Sep 2015 13:19
 Operator : IZ/UA
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 14:50:36 2015
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP090315.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 14:50:01 2015
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.170	4.190	2193360	2947800	49.929	50.912
2)	SA Decachlor...	11.500	9.289	1828665	2670315	48.022	47.307
Target Compounds							
4)	L1 AR-1016-2	0.000	5.511	0	643828	N.D.	436.543 #
5)	L1 AR-1016-3	0.000	5.511	0	643828	N.D.	555.337 #
6)	L1 AR-1016-4	6.842	5.553	255370	160583	258.394	108.331 #
7)	L1 AR-1016-5	6.986	5.729	107949	345047	101.803	222.726 #
13)	L3 AR-1232-3	0.000	5.511	0	643828	N.D.	1089.247 #
14)	L3 AR-1232-4	0.000	5.729	0	345047	N.D.	576.736 #
15)	L3 AR-1232-5	6.986	6.079	107949	1665885	255.247	2452.160 #
18)	L4 AR-1242-3	0.000	5.729	0	345047	N.D.	259.345 #
19)	L4 AR-1242-4	7.245	6.079	339166	1665885	379.211	1075.935 #
20)	L4 AR-1242-5	7.289	6.119	253073	185835	270.933	140.888 #
23)	L5 AR-1248-3	6.842	5.511	255370	643828	189.093	382.548 #
24)	L5 AR-1248-4	7.245	5.553	339166	160583	215.445	91.221 #
25)	L5 AR-1248-5	7.289	6.079	253073	1665885	165.141	567.449 #
26)	L6 AR-1254-1	7.220	6.079	866687	1665885	500.000	500.000
27)	L6 AR-1254-2	7.436	6.221	1258156	1442477	500.000	500.000
28)	L6 AR-1254-3	7.809	6.539	1347728	1178900	500.000	500.000
29)	L6 AR-1254-4	8.094	6.629	1019941	2392546	500.000	500.000
30)	L6 AR-1254-5	8.542	7.268	1266999	2078112	500.000	500.000
31)	L7 AR-1260-1	7.970	6.755	568723	1112358	301.320	354.213
32)	L7 AR-1260-2	8.228	6.937	611597	886913	265.862	229.403
33)	L7 AR-1260-3	8.614	7.096	134847	931907	76.598	265.146 #
34)	L7 AR-1260-4	8.881	7.785	426566	538818	211.494	78.003 #
35)	L7 AR-1260-5	9.264	8.148	189620	242712	46.531	47.951



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD P\Data\PP090315\
Data File : PP011538.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Sep 2015 13:19
Operator : IZ/UA
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 14:50:36 2015
Quant Method : P:\HPCHEM1\ECD_P\methods\PP090315.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 14:50:01 2015
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 2193360
Conc: 49.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.190 min
Delta R.T.: 0.000 min
Response: 2947800
Conc: 50.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.500 min
Delta R.T.: 0.000 min
Response: 1828665
Conc: 48.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.289 min
Delta R.T.: 0.000 min
Response: 2670315
Conc: 47.31 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T.: 6.446 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#4 AR-1016-2

R.T.: 5.511 min
Delta R.T.: 0.037 min
Response: 643828
Conc: 436.54 ng/ml




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T.: 6.548 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 643828
Conc: 555.34 ng/ml



#6 AR-1016-4


R.T.: 6.842 min
Delta R.T.: 0.000 min
Response: 255370
Conc: 258.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#6 AR-1016-4

R.T.: 5.553 min
Delta R.T.: -0.002 min
Response: 160583
Conc: 108.33 ng/ml




#7 AR-1016-5

R.T.: 6.986 min
Delta R.T.: 0.001 min
Response: 107949
Conc: 101.80 ng/ml



#7 AR-1016-5


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 345047
Conc: 222.73 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 6.446 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#13 AR-1232-3

R.T.: 5.511 min
Delta R.T.: 0.034 min
Response: 643828
Conc: 1089.25 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T.: 6.549 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 345047
Conc: 576.74 ng/ml



#15 AR-1232-5


R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 107949
Conc: 255.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#15 AR-1232-5

R.T.: 6.079 min
Delta R.T.: 0.001 min
Response: 1665885
Conc: 2452.16 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 6.446 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 345047
Conc: 259.34 ng/ml



#19 AR-1242-4


R.T.: 7.245 min
Delta R.T.: 0.000 min
Response: 339166
Conc: 379.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#19 AR-1242-4

R.T.: 6.079 min
Delta R.T.: 0.002 min
Response: 1665885
Conc: 1075.93 ng/ml




#20 AR-1242-5

R.T.: 7.289 min
Delta R.T.: 0.002 min
Response: 253073
Conc: 270.93 ng/ml



#20 AR-1242-5


R.T.: 6.119 min
Delta R.T.: 0.000 min
Response: 185835
Conc: 140.89 ng/ml



#23 AR-1248-3


R.T.: 6.842 min
Delta R.T.: 0.001 min
Response: 255370
Conc: 189.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#23 AR-1248-3

R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 643828
Conc: 382.55 ng/ml




#24 AR-1248-4

R.T.: 7.245 min
Delta R.T.: -0.001 min
Response: 339166
Conc: 215.45 ng/ml



#24 AR-1248-4


R.T.: 5.553 min
Delta R.T.: -0.002 min
Response: 160583
Conc: 91.22 ng/ml



#25 AR-1248-5


R.T.: 7.289 min
Delta R.T.: 0.001 min
Response: 253073
Conc: 165.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#25 AR-1248-5

R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 1665885
Conc: 567.45 ng/ml




#26 AR-1254-1

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 866687
Conc: 500.00 ng/ml



#26 AR-1254-1


R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 1665885
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 7.436 min
Delta R.T.: 0.000 min
Response: 1258156
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#27 AR-1254-2

R.T.: 6.221 min
Delta R.T.: 0.000 min
Response: 1442477
Conc: 500.00 ng/ml




#28 AR-1254-3

R.T.: 7.809 min
Delta R.T.: 0.000 min
Response: 1347728
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.539 min
Delta R.T.: 0.000 min
Response: 1178900
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 8.094 min
Delta R.T.: 0.000 min
Response: 1019941
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#29 AR-1254-4

R.T.: 6.629 min
Delta R.T.: 0.000 min
Response: 2392546
Conc: 500.00 ng/ml




#30 AR-1254-5

R.T.: 8.542 min
Delta R.T.: 0.000 min
Response: 1266999
Conc: 500.00 ng/ml



#30 AR-1254-5


R.T.: 7.268 min
Delta R.T.: 0.000 min
Response: 2078112
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 7.970 min
Delta R.T.: 0.000 min
Response: 568723
Conc: 301.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#31 AR-1260-1

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 1112358
Conc: 354.21 ng/ml




#32 AR-1260-2

R.T.: 8.228 min
Delta R.T.: 0.000 min
Response: 611597
Conc: 265.86 ng/ml



#32 AR-1260-2


R.T.: 6.937 min
Delta R.T.: 0.000 min
Response: 886913
Conc: 229.40 ng/ml



#33 AR-1260-3


R.T.: 8.614 min
Delta R.T.: -0.005 min
Response: 134847
Conc: 76.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#33 AR-1260-3

R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 931907
Conc: 265.15 ng/ml




#34 AR-1260-4

R.T.: 8.881 min
Delta R.T.: -0.005 min
Response: 426566
Conc: 211.49 ng/ml



#34 AR-1260-4


R.T.: 7.785 min
Delta R.T.: -0.013 min
Response: 538818
Conc: 78.00 ng/ml



#35 AR-1260-5

R.T.: 9.264 min
Delta R.T.: 0.000 min
Response: 189620
Conc: 46.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#35 AR-1260-5

R.T.: 8.148 min
Delta R.T.: 0.000 min
Response: 242712
Conc: 47.95 ng/ml