

Method Path : P:\HPCHEM1\ECD P\methods\
 Method File : PP090315.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 03 14:57:56 2015
 Response Via : Initial Calibration

Calibration Files

500 =PP011540.D 1000 =PP011529.D 750 =PP011530.D
 250 =PP011532.D 50 =PP011533.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		4.209	4.205	4.442	4.431	4.499	4.357	E4	3.20
2) SA	Decachlorobiphenyl		5.532	3.484	3.625	3.949	4.324	4.183	E4	19.62
3) L1	AR-1016-1		0.997	0.959	1.016	1.071	1.052	1.019	E3	4.35
4) L1	AR-1016-2		1.230	1.182	1.249	1.329	1.428	1.284	E3	7.52
5) L1	AR-1016-3		1.008	0.966	1.017	1.087	1.106	1.037	E3	5.58
6) L1	AR-1016-4		0.960	0.926	0.972	1.026	1.058	0.988	E3	5.39
7) L1	AR-1016-5		1.022	0.983	1.026	1.054	1.217	1.060	E3	8.61
8) L2	AR-1221-1		4.309					4.309	E2	0.00
9) L2	AR-1221-2		3.330					3.330	E2	0.00
10) L2	AR-1221-3		1.062					1.062	E3	0.00
11) L3	AR-1232-1		8.745					8.745	E2	0.00
12) L3	AR-1232-2		4.269					4.269	E2	0.00
13) L3	AR-1232-3		5.370					5.370	E2	0.00
14) L3	AR-1232-4		4.274					4.274	E2	0.00
15) L3	AR-1232-5		4.229					4.229	E2	0.00
16) L4	AR-1242-1		9.067					9.067	E2	0.00
17) L4	AR-1242-2		8.643					8.643	E2	0.00
18) L4	AR-1242-3		1.065					1.065	E3	0.00
19) L4	AR-1242-4		8.944					8.944	E2	0.00
20) L4	AR-1242-5		9.341					9.341	E2	0.00
21) L5	AR-1248-1		8.948					8.948	E2	0.00
22) L5	AR-1248-2		1.021					1.021	E3	0.00
23) L5	AR-1248-3		1.350					1.350	E3	0.00
24) L5	AR-1248-4		1.574					1.574	E3	0.00
25) L5	AR-1248-5		1.532					1.532	E3	0.00
26) L6	AR-1254-1		1.733					1.733	E3	0.00
27) L6	AR-1254-2		2.516					2.516	E3	0.00
28) L6	AR-1254-3		2.695					2.695	E3	0.00
29) L6	AR-1254-4		2.040					2.040	E3	0.00
30) L6	AR-1254-5		2.534					2.534	E3	0.00
31) L7	AR-1260-1		1.855	1.770	1.845	1.940	2.027	1.887	E3	5.23
32) L7	AR-1260-2		2.266	2.156	2.235	2.333	2.513	2.300	E3	5.84
33) L7	AR-1260-3		1.731	1.649	1.712	1.817	1.894	1.760	E3	5.45
34) L7	AR-1260-4		2.020	1.919	1.983	2.125	2.037	2.017	E3	3.75
35) L7	AR-1260-5		3.972	3.808	3.910	4.108	4.577	4.075	E3	7.38
36) L8	AR-1262-1		1.850					1.850	E3	0.00
37) L8	AR-1262-2		2.226					2.226	E3	0.00
38) L8	AR-1262-3		3.101					3.101	E3	0.00
39) L8	AR-1262-4		2.964					2.964	E3	0.00
40) L8	AR-1262-5		5.610					5.610	E3	0.00
41) L9	AR-1268-1		5.319					5.319	E3	0.00
42) L9	AR-1268-2		5.407					5.407	E3	0.00
43) L9	AR-1268-3		3.924					3.924	E3	0.00
44) L9	AR-1268-4		1.696					1.696	E3	0.00
45) L9	AR-1268-5		1.132					1.132	E4	0.00

Signal #2 Calibration Files

500 =PP011540.D 1000 =PP011529.D 750 =PP011530.D
 250 =PP011532.D 50 =PP011533.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		5.644	5.553	5.869	5.790	5.842	5.740	E4	2.37
2) SA	Decachlorobiphenyl		8.212	5.179	5.366	5.840	6.498	6.219	E4	19.70

Method Path : P:\HPCHEM1\ECD P\methods\
 Method File : PP090315.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 03 14:57:56 2015
 Response Via : Initial Calibration

Calibration Files

500 =PP011540.D 1000 =PP011529.D 750 =PP011530.D
 250 =PP011532.D 50 =PP011533.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	1.422	1.359	1.370	1.453	1.557	1.432	E3	5.57
4)	L1	AR-1016-2	1.417	1.387	1.445	1.524	1.602	1.475	E3	5.91
5)	L1	AR-1016-3	1.116	1.066	1.120	1.220	1.275	1.159	E3	7.37
6)	L1	AR-1016-4	1.407	1.344	1.417	1.533	1.710	1.482	E3	9.75
7)	L1	AR-1016-5	1.489	1.438	1.503	1.593	1.724	1.549	E3	7.27
8)	L2	AR-1221-1	5.742					5.742	E2	0.00
9)	L2	AR-1221-2	4.250					4.250	E2	0.00
10)	L2	AR-1221-3	1.342					1.342	E3	0.00
11)	L3	AR-1232-1	1.134					1.134	E3	0.00
12)	L3	AR-1232-2	5.881					5.881	E2	0.00
13)	L3	AR-1232-3	5.911					5.911	E2	0.00
14)	L3	AR-1232-4	5.983					5.983	E2	0.00
15)	L3	AR-1232-5	6.794					6.794	E2	0.00
16)	L4	AR-1242-1	1.172					1.172	E3	0.00
17)	L4	AR-1242-2	1.664					1.664	E3	0.00
18)	L4	AR-1242-3	1.330					1.330	E3	0.00
19)	L4	AR-1242-4	1.548					1.548	E3	0.00
20)	L4	AR-1242-5	1.319					1.319	E3	0.00
21)	L5	AR-1248-1	1.180					1.180	E3	0.00
22)	L5	AR-1248-2	1.180					1.180	E3	0.00
23)	L5	AR-1248-3	1.683					1.683	E3	0.00
24)	L5	AR-1248-4	1.760					1.760	E3	0.00
25)	L5	AR-1248-5	2.936					2.936	E3	0.00
26)	L6	AR-1254-1	3.332					3.332	E3	0.00
27)	L6	AR-1254-2	2.885					2.885	E3	0.00
28)	L6	AR-1254-3	2.358					2.358	E3	0.00
29)	L6	AR-1254-4	4.785					4.785	E3	0.00
30)	L6	AR-1254-5	4.156					4.156	E3	0.00
31)	L7	AR-1260-1	3.004	2.875	2.988	3.213	3.622	3.140	E3	9.42
32)	L7	AR-1260-2	3.697	3.537	3.677	3.961	4.459	3.866	E3	9.44
33)	L7	AR-1260-3	3.408	3.283	3.381	3.603	3.898	3.515	E3	6.94
34)	L7	AR-1260-4	6.777	6.561	6.777	7.046	7.376	6.908	E3	4.54
35)	L7	AR-1260-5	4.909	4.721	4.871	5.144	5.663	5.062	E3	7.29
36)	L8	AR-1262-1	3.619					3.619	E3	0.00
37)	L8	AR-1262-2	5.270					5.270	E3	0.00
38)	L8	AR-1262-3	4.407					4.407	E3	0.00
39)	L8	AR-1262-4	9.348					9.348	E3	0.00
40)	L8	AR-1262-5	6.881					6.881	E3	0.00
41)	L9	AR-1268-1	8.565					8.565	E3	0.00
42)	L9	AR-1268-2	8.612					8.612	E3	0.00
43)	L9	AR-1268-3	6.151					6.151	E3	0.00
44)	L9	AR-1268-4	2.601					2.601	E3	0.00
45)	L9	AR-1268-5	1.761					1.761	E4	0.00

(#) = Out of Range