

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP090319.M
 Title : 8080.M
 Last Update : Wed Sep 04 06:57:46 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP030416.D 1000 =PP030417.D 1500 =PP030418.D
 500 =PP030415.D 200 =PP030414.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	4.617	4.438	4.429	4.667	4.646	4.559	E5	2.54
2) T	3,5-DICHLOROBENZOI...	4.766	4.562	4.368	4.884	4.944	4.705	E5	5.06
3) T	4-Nitrophenol	3.049	2.989	2.887	3.093	2.898	2.983	E5	3.05
4) S	2,4-DCAA	2.827	2.733	2.584	2.908	2.950	2.800	E5	5.24
5) T	DICAMBA	1.173	1.137	1.094	1.192	1.197	1.159	E6	3.72
6) T	MCPP	7.044	6.887	6.659	6.947	6.235	6.754	E5	4.78
7) T	MCPA	0.996	0.982	0.942	1.003	0.941	0.973	E6	3.05
8) T	DICHLORPROP	2.800	2.710	2.546	2.893	2.999	2.789	E5	6.22
9) T	2,4-D	3.044	3.006	2.836	3.095	3.073	3.011	E5	3.42
10) T	Pentachlorophenol	4.565	4.448	4.178	4.625	4.505	4.464	E6	3.88
11) T	2,4,5-TP (SILVEX)	1.536	1.521	1.445	1.539	1.509	1.510	E6	2.53
12) T	2,4,5-T	1.391	1.410	1.359	1.373	1.263	1.359	E6	4.18
13) T	2,4-DB	1.893	1.871	1.783	1.917	1.898	1.872	E5	2.81
14) T	DINOSEB	8.778	9.054	9.015	8.366	7.293	8.501	E5	8.57
15) T	Picloram	1.630	1.612	1.653	1.574	1.419	1.578	E6	5.92
16) T	DCPA	1.924	1.923	1.844	1.907	1.842	1.888	E6	2.21

Signal #2 Calibration Files

750 =PP030416.D 1000 =PP030417.D 1500 =PP030418.D
 500 =PP030415.D 200 =PP030414.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	2.831	2.710	2.727	2.788	2.849	2.781	E5	2.20
2) T	3,5-DICHLOROBENZOI...	2.663	2.527	2.477	2.646	2.725	2.608	E5	3.93
3) T	4-Nitrophenol	1.108	1.083	1.069	1.038	0.795	1.018	E5	12.50
4) S	2,4-DCAA	1.995	1.915	1.871	1.973	2.022	1.955	E5	3.14
5) T	DICAMBA	7.642	7.374	7.257	7.496	7.595	7.473	E5	2.12
6) T	MCPP	3.923	3.890	3.951	3.652	2.716	3.626	E5	14.41
7) T	MCPA	6.548	6.467	6.376	6.364	5.965	6.344	E5	3.54
8) T	DICHLORPROP	1.826	1.770	1.706	1.820	1.879	1.800	E5	3.63
9) T	2,4-D	2.050	2.022	1.961	1.943	1.905	1.976	E5	2.99
10) T	Pentachlorophenol	2.969	2.854	2.715	2.948	2.952	2.888	E6	3.68
11) T	2,4,5-TP (SILVEX)	1.042	1.024	0.993	1.025	1.024	1.021	E6	1.74
12) T	2,4,5-T	8.680	8.816	8.788	8.263	7.506	8.411	E5	6.57
13) T	2,4-DB	1.229	1.210	1.183	1.215	1.181	1.204	E5	1.74
14) T	DINOSEB	5.692	5.820	5.883	5.275	4.684	5.471	E5	9.13
15) T	Picloram	1.202	1.191	1.237	1.140	0.989	1.152	E6	8.46
16) T	DCPA	1.322	1.303	1.254	1.303	1.325	1.302	E6	2.19

(#) = Out of Range