

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090324\
 Data File : PP067076.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2024 08:30
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 09:42:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 04 09:40:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.620	3.911	45976817	42606702	72.125	73.608
2) SA Decachlor...	10.470	9.041	64176179	69629174	73.173	73.025
Target Compounds						
21) L5 AR-1248-1	5.791	5.017	9392169	9154459	706.972	723.554
22) L5 AR-1248-2	6.064	5.256	16893240	16742845	708.232	690.769
23) L5 AR-1248-3	6.269	5.300	19229310	17494831	713.030	689.356
24) L5 AR-1248-4	6.673	5.475	19147868	19883978	692.312	713.731
25) L5 AR-1248-5	6.712	5.873	21408346	17481669	686.462	692.078



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090324\
Data File : PP067076.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2024 08:30
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 09:42:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 04 09:40:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.620 min
Delta R.T.: 0.000 min
Response: 45976817
Conc: 72.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.911 min
Delta R.T.: 0.000 min
Response: 42606702
Conc: 73.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.470 min
Delta R.T.: 0.002 min
Response: 64176179
Conc: 73.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.041 min
Delta R.T.: 0.002 min
Response: 69629174
Conc: 73.03 ng/ml



#21 AR-1248-1


R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 9392169
Conc: 706.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 9154459
Conc: 723.55 ng/ml




#22 AR-1248-2

R.T.: 6.064 min
Delta R.T.: 0.000 min
Response: 16893240
Conc: 708.23 ng/ml



#22 AR-1248-2


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 16742845
Conc: 690.77 ng/ml



#23 AR-1248-3


R.T.: 6.269 min
Delta R.T.: 0.000 min
Response: 19229310
Conc: 713.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.300 min
Delta R.T.: 0.000 min
Response: 17494831
Conc: 689.36 ng/ml




#24 AR-1248-4

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 19147868
Conc: 692.31 ng/ml



#24 AR-1248-4


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 19883978
Conc: 713.73 ng/ml



#25 AR-1248-5

R.T.: 6.712 min
Delta R.T.: 0.000 min
Response: 21408346
Conc: 686.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.873 min
Delta R.T.: 0.000 min
Response: 17481669
Conc: 692.08 ng/ml