

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090324\
 Data File : PP067077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2024 08:46
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 09:40:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 04 09:40:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.620	3.911	31873142	28941654	50.000	50.000
2) SA Decachlor...	10.468	9.039	43852338	47674594	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.790	5.017	6642537	6326040	500.000	500.000
22) L5 AR-1248-2	6.063	5.256	11926353	12118997	500.000	500.000
23) L5 AR-1248-3	6.269	5.300	13484219	12689259	500.000	500.000
24) L5 AR-1248-4	6.672	5.475	13828927	13929600	500.000	500.000
25) L5 AR-1248-5	6.712	5.873	15593238	12629845	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090324\
Data File : PP067077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2024 08:46
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 09:40:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090324.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 04 09:40:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.620 min
Delta R.T.: 0.000 min
Response: 31873142
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.911 min
Delta R.T.: 0.000 min
Response: 28941654
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.468 min
Delta R.T.: 0.000 min
Response: 43852338
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.039 min
Delta R.T.: 0.000 min
Response: 47674594
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.790 min
Delta R.T.: 0.000 min
Response: 6642537
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 6326040
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 11926353
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 12118997
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.269 min
Delta R.T.: 0.000 min
Response: 13484219
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.300 min
Delta R.T.: 0.000 min
Response: 12689259
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.672 min
Delta R.T.: 0.000 min
Response: 13828927
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 13929600
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.712 min
Delta R.T.: 0.000 min
Response: 15593238
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.873 min
Delta R.T.: 0.000 min
Response: 12629845
Conc: 500.00 ng/ml