

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
 Data File : PP059761.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 12:57
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 14:05:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 06 13:59:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.608	144.4E6	194.7E6	97.384	97.836
2) SA Decachlor...	10.155	8.628	162.9E6	211.4E6	97.795	97.696
Target Compounds						
41) L9 AR-1268-1	8.646	7.529	129.0E6	205.3E6	977.083	978.737
42) L9 AR-1268-2	8.740	7.594	114.5E6	184.8E6	977.748	980.207
43) L9 AR-1268-3	8.969	7.799	105.7E6	159.5E6	961.257	977.761
44) L9 AR-1268-4	9.396	8.090	38268132	60357490	981.781	980.227
45) L9 AR-1268-5	9.813	8.377	317.5E6	425.1E6	987.896	989.424



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
Data File : PP059761.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 12:57
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 14:05:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 06 13:59:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: 0.000 min
Response: 144383654
Conc: 97.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.608 min
Delta R.T.: 0.000 min
Response: 194746277
Conc: 97.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.155 min
Delta R.T.: 0.000 min
Response: 162860666
Conc: 97.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: 0.000 min
Response: 211407994
Conc: 97.70 ng/ml



#41 AR-1268-1


R.T.: 8.646 min
Delta R.T.: 0.000 min
Response: 129012228
Conc: 977.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.529 min
Delta R.T.: 0.000 min
Response: 205344751
Conc: 978.74 ng/ml




#42 AR-1268-2

R.T.: 8.740 min
Delta R.T.: 0.000 min
Response: 114452492
Conc: 977.75 ng/ml



#42 AR-1268-2


R.T.: 7.594 min
Delta R.T.: 0.000 min
Response: 184840914
Conc: 980.21 ng/ml



#43 AR-1268-3


R.T.: 8.969 min
Delta R.T.: 0.000 min
Response: 105711494
Conc: 961.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.799 min
Delta R.T.: 0.000 min
Response: 159522331
Conc: 977.76 ng/ml




#44 AR-1268-4

R.T.: 9.396 min
Delta R.T.: 0.000 min
Response: 38268132
Conc: 981.78 ng/ml



#44 AR-1268-4


R.T.: 8.090 min
Delta R.T.: 0.000 min
Response: 60357490
Conc: 980.23 ng/ml



#45 AR-1268-5

R.T.: 9.813 min
Delta R.T.: 0.000 min
Response: 317514690
Conc: 987.90 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.377 min
Delta R.T.: 0.000 min
Response: 425111714
Conc: 989.42 ng/ml