

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
 Data File : PP059746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 08:53
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 10:41:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 06 10:40:20 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.608	108.6E6	147.0E6	73.074	73.665
2) SA Decachlor...	10.155	8.628	77796074	100.4E6	74.762	72.689
Target Compounds						
16) L4 AR-1242-1	5.559	4.696	29213478	36806867	704.378	713.859
17) L4 AR-1242-2	5.581	4.715	43066708	54596185	714.838	725.059
18) L4 AR-1242-3	5.644	4.890	27681568	29306040	719.658	724.438
19) L4 AR-1242-4	5.743	4.976	22568501	30076605	710.679	711.879
20) L4 AR-1242-5	6.483	5.501	20710829	34173771	718.321	719.816

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
Data File : PP059746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 08:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 10:41:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 06 10:40:20 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

