

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
 Data File : PP059744.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 08:20
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 10:38:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 06 10:38:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.608	72703246	98373294	50.000	50.000
2) SA Decachlor...	10.155	8.628	50584704	66617446	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.755	3.983	16408416	22045643	500.000	500.000
12) L3 AR-1232-2	5.288	4.715	9207274	19844115	500.000	500.000
13) L3 AR-1232-3	5.580	4.890	15805065	10240159	500.000	500.000
14) L3 AR-1232-4	5.742	4.975	8085936	10002500	500.000	500.000
15) L3 AR-1232-5	5.879	5.146	6728174	10803366	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090623\
Data File : PP059744.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 08:20
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 10:38:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 06 10:38:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: 0.000 min
Response: 72703246
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.608 min
Delta R.T.: 0.000 min
Response: 98373294
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.155 min
Delta R.T.: 0.000 min
Response: 50584704
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: 0.000 min
Response: 66617446
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.755 min
Delta R.T.: 0.000 min
Response: 16408416
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 22045643
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 9207274
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.715 min
Delta R.T.: 0.000 min
Response: 19844115
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 5.580 min
Delta R.T.: 0.000 min
Response: 15805065
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 4.890 min
Delta R.T.: 0.000 min
Response: 10240159
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 8085936
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 10002500
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.879 min
Delta R.T.: 0.000 min
Response: 6728174
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.146 min
Delta R.T.: 0.000 min
Response: 10803366
Conc: 500.00 ng/ml