

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP090623\  
 Data File : PP059792.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 06 Sep 2023 21:35  
 Operator : YP\AJ  
 Sample : PB155318BSD  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 PB155318BSD

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Sep 07 00:28:39 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP090623.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Sep 06 14:24:13 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|--------|-------|----------|----------|--------|--------|
| -----                       |        |       |          |          |        |        |
| System Monitoring Compounds |        |       |          |          |        |        |
| 1) SA Tetrachlo...          | 4.382  | 3.607 | 26322860 | 33391559 | 17.856 | 16.905 |
| 2) SA Decachlor...          | 10.156 | 8.626 | 20745944 | 26648935 | 20.208 | 19.522 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP090623\  
Data File : PP059792.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 06 Sep 2023 21:35  
Operator : YP\AJ  
Sample : PB155318BSD  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
PB155318BSD

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 07 00:28:39 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP090623.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Sep 06 14:24:13 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min  
Delta R.T.: 0.001 min  
Response: 26322860  
Conc: 17.86 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
PB155318BSD




#1 Tetrachloro-m-xylene

R.T.: 3.607 min  
Delta R.T.: 0.000 min  
Response: 33391559  
Conc: 16.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.156 min  
Delta R.T.: 0.003 min  
Response: 20745944  
Conc: 20.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.626 min  
Delta R.T.: -0.002 min  
Response: 26648935  
Conc: 19.52 ng/ml