

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
 Data File : PP039182.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Sep 2021 12:35
 Operator : AJ\MA
 Sample : M3650-04 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 08:46:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.871	3.857	147370	71422	3.847	2.978
2) SA Decachlor...	10.841	9.109	137957	80317	4.191	3.670
Target Compounds						
6) L1 AR-1016-4	6.394	0.000	14249	0	14.123	N.D. #
9) L2 AR-1221-2	5.210	0.000	12661	0	34.094	N.D. #
14) L3 AR-1232-4	6.394	0.000	14249	0	30.333	N.D. #
15) L3 AR-1232-5	6.499f	0.000	28836	0	87.098	N.D. #
19) L4 AR-1242-4	6.394	0.000	14249	0	18.511	N.D. #
22) L5 AR-1248-2	6.499f	0.000	28836	0	27.027	N.D. #
28) L6 AR-1254-3	7.733f	0.000	7289	0	2.978	N.D. #
30) L6 AR-1254-5	8.432	0.000	13123	0	6.580	N.D. #
32) L7 AR-1260-2	8.117f	0.000	4614	0	2.323	N.D. #
33) L7 AR-1260-3	8.514	0.000	12767	0	9.458	N.D. #
34) L7 AR-1260-4	8.755f	0.000	13578	0	8.423	N.D. #
35) L7 AR-1260-5	9.058	0.000	17191	0	6.206	N.D. #
36) L8 AR-1262-1	8.514	0.000	12767	0	5.731	N.D. #
37) L8 AR-1262-2	9.058	0.000	17191	0	4.643	N.D. #
38) L8 AR-1262-3	9.384f	0.000	18215	0	9.833	N.D. #
41) L9 AR-1268-1	9.384f	0.000	18215	0	3.713	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
Data File : PP039182.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Sep 2021 12:35
Operator : AJ\MA
Sample : M3650-04 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 08:46:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 04:15:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.871 min
Delta R.T.: 0.000 min
Response: 147370
Conc: 3.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.857 min
Delta R.T.: 0.000 min
Response: 71422
Conc: 2.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.841 min
Delta R.T.: -0.002 min
Response: 137957
Conc: 4.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.109 min
Delta R.T.: -0.003 min
Response: 80317
Conc: 3.67 ng/ml



#6 AR-1016-4


R.T.: 6.394 min
Delta R.T.: 0.006 min
Response: 14249
Conc: 14.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.362 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.210 min
Delta R.T.: -0.010 min
Response: 12661
Conc: 34.09 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.209 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 6.394 min
Delta R.T.: 0.006 min
Response: 14249
Conc: 30.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 5.405 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 6.499 min
Delta R.T.: 0.015 min
Response: 28836
Conc: 87.10 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.588 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 6.394 min
Delta R.T.: 0.007 min
Response: 14249
Conc: 18.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.405 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 6.499 min
Delta R.T.: 0.016 min
Response: 28836
Conc: 27.03 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 5.361 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.733 min
Delta R.T.: 0.024 min
Response: 7289
Conc: 2.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.432 min
Delta R.T.: 0.000 min
Response: 13123
Conc: 6.58 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.209 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.117 min
Delta R.T.: -0.026 min
Response: 4614
Conc: 2.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.879 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.514 min
Delta R.T.: 0.006 min
Response: 12767
Conc: 9.46 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.030 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.755 min
Delta R.T.: 0.016 min
Response: 13578
Conc: 8.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.512 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.058 min
Delta R.T.: -0.003 min
Response: 17191
Conc: 6.21 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 7.763 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 8.514 min
Delta R.T.: 0.010 min
Response: 12767
Conc: 5.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.208 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 9.058 min
Delta R.T.: 0.001 min
Response: 17191
Conc: 4.64 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.760 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 9.384 min
Delta R.T.: 0.024 min
Response: 18215
Conc: 9.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.044 min
Response: 0
Conc: N.D.



#41 AR-1268-1

R.T.: 9.384 min
Delta R.T.: 0.023 min
Response: 18215
Conc: 3.71 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 8.044 min
Response: 0
Conc: N.D.