

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
 Data File : PP039203.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Sep 2021 20:03
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 09:00:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.870	3.855	1996903	1224823	52.123	51.064
2)	SA Decachlor...	10.837	9.107	1888336	1182689	57.368	54.041
Target Compounds							
26)	L6 AR-1254-1	7.096	5.963	798703	650173	523.818	516.929
27)	L6 AR-1254-2	7.325	6.123	1186471	584594	504.597	506.014
28)	L6 AR-1254-3	7.707	6.539	1266614	924903	517.516	522.068
29)	L6 AR-1254-4	8.004	6.780	967307	583833	582.782	638.884
30)	L6 AR-1254-5	8.431	7.207	1071000	890717	537.014	538.558



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
Data File : PP039203.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Sep 2021 20:03
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 09:00:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 04:15:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.870 min
Delta R.T.: -0.001 min
Response: 1996903
Conc: 52.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.855 min
Delta R.T.: 0.000 min
Response: 1224823
Conc: 51.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.837 min
Delta R.T.: -0.006 min
Response: 1888336
Conc: 57.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.107 min
Delta R.T.: -0.005 min
Response: 1182689
Conc: 54.04 ng/ml



#26 AR-1254-1


R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 798703
Conc: 523.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.963 min
Delta R.T.: -0.001 min
Response: 650173
Conc: 516.93 ng/ml




#27 AR-1254-2

R.T.: 7.325 min
Delta R.T.: 0.000 min
Response: 1186471
Conc: 504.60 ng/ml



#27 AR-1254-2


R.T.: 6.123 min
Delta R.T.: -0.002 min
Response: 584594
Conc: 506.01 ng/ml



#28 AR-1254-3


R.T.: 7.707 min
Delta R.T.: -0.002 min
Response: 1266614
Conc: 517.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.539 min
Delta R.T.: -0.002 min
Response: 924903
Conc: 522.07 ng/ml




#29 AR-1254-4

R.T.: 8.004 min
Delta R.T.: -0.002 min
Response: 967307
Conc: 582.78 ng/ml



#29 AR-1254-4


R.T.: 6.780 min
Delta R.T.: -0.004 min
Response: 583833
Conc: 638.88 ng/ml



#30 AR-1254-5

R.T.: 8.431 min
Delta R.T.: -0.001 min
Response: 1071000
Conc: 537.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.207 min
Delta R.T.: -0.002 min
Response: 890717
Conc: 538.56 ng/ml