

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
 Data File : PP039217.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Sep 2021 00:33
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 09:09:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 04:15:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.869	3.855	2030642	1223254	53.004	50.998
2) SA Decachlor...	10.835	9.105	1900616	1181438	57.741	53.984
Target Compounds						
21) L5 AR-1248-1	6.188	5.097	429019	232481	559.619	552.036
22) L5 AR-1248-2	6.481	5.358	555240	329948	520.403	492.404
23) L5 AR-1248-3	6.697	5.402	663697	332783	519.223	498.359
24) L5 AR-1248-4	7.127	5.584	741056	379411	551.841	476.212
25) L5 AR-1248-5	7.166	6.005	756467	397099	539.026	570.101



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP090721\
Data File : PP039217.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Sep 2021 00:33
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 09:09:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 04:15:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: -0.002 min
Response: 2030642
Conc: 53.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.855 min
Delta R.T.: -0.001 min
Response: 1223254
Conc: 51.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.835 min
Delta R.T.: -0.008 min
Response: 1900616
Conc: 57.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.105 min
Delta R.T.: -0.006 min
Response: 1181438
Conc: 53.98 ng/ml



#21 AR-1248-1


R.T.: 6.188 min
Delta R.T.: -0.003 min
Response: 429019
Conc: 559.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.097 min
Delta R.T.: -0.002 min
Response: 232481
Conc: 552.04 ng/ml




#22 AR-1248-2

R.T.: 6.481 min
Delta R.T.: -0.003 min
Response: 555240
Conc: 520.40 ng/ml



#22 AR-1248-2


R.T.: 5.358 min
Delta R.T.: -0.002 min
Response: 329948
Conc: 492.40 ng/ml



#23 AR-1248-3


R.T.: 6.697 min
Delta R.T.: -0.002 min
Response: 663697
Conc: 519.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.402 min
Delta R.T.: -0.003 min
Response: 332783
Conc: 498.36 ng/ml




#24 AR-1248-4

R.T.: 7.127 min
Delta R.T.: -0.003 min
Response: 741056
Conc: 551.84 ng/ml



#24 AR-1248-4


R.T.: 5.584 min
Delta R.T.: -0.003 min
Response: 379411
Conc: 476.21 ng/ml



#25 AR-1248-5

R.T.: 7.166 min
Delta R.T.: -0.004 min
Response: 756467
Conc: 539.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.005 min
Delta R.T.: -0.005 min
Response: 397099
Conc: 570.10 ng/ml