

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091123\
 Data File : PP059831.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2023 22:40
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 01:32:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 12 01:31:42 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.381	3.605	41710740	58450732	26.256	25.256
2) SA Decachlor...	10.153	8.620	32694518	48186079	26.855	26.837
Target Compounds						
26) L6 AR-1254-1	6.417	5.496	14423535	31126962	275.817	268.130
27) L6 AR-1254-2	6.637	5.645	20499600	27554139	272.455	269.189
28) L6 AR-1254-3	7.004	6.048	20397908	41468785	271.278	264.417
29) L6 AR-1254-4	7.291	6.280	13849326	24373565	267.853	265.535
30) L6 AR-1254-5	7.712	6.698	15761064	37053340	266.679	264.422



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091123\
Data File : PP059831.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2023 22:40
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 01:32:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 12 01:31:42 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.381 min
Delta R.T.: 0.000 min
Response: 41710740
Conc: 26.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 58450732
Conc: 25.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.153 min
Delta R.T.: 0.000 min
Response: 32694518
Conc: 26.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.620 min
Delta R.T.: 0.000 min
Response: 48186079
Conc: 26.84 ng/ml



#26 AR-1254-1


R.T.: 6.417 min
Delta R.T.: 0.000 min
Response: 14423535
Conc: 275.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.496 min
Delta R.T.: 0.000 min
Response: 31126962
Conc: 268.13 ng/ml




#27 AR-1254-2

R.T.: 6.637 min
Delta R.T.: 0.000 min
Response: 20499600
Conc: 272.46 ng/ml



#27 AR-1254-2


R.T.: 5.645 min
Delta R.T.: 0.000 min
Response: 27554139
Conc: 269.19 ng/ml



#28 AR-1254-3


R.T.: 7.004 min
Delta R.T.: 0.000 min
Response: 20397908
Conc: 271.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.048 min
Delta R.T.: 0.000 min
Response: 41468785
Conc: 264.42 ng/ml




#29 AR-1254-4

R.T.: 7.291 min
Delta R.T.: 0.000 min
Response: 13849326
Conc: 267.85 ng/ml



#29 AR-1254-4


R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 24373565
Conc: 265.54 ng/ml



#30 AR-1254-5

R.T.: 7.712 min
Delta R.T.: 0.000 min
Response: 15761064
Conc: 266.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.698 min
Delta R.T.: 0.000 min
Response: 37053340
Conc: 264.42 ng/ml