

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091222\
 Data File : PP051260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2022 10:19
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 11:56:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 11:55:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.457	3.701	88749532	63569531	50.000	50.000
2) SA Decachlor...	10.310	8.812	54809351	45335771	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.635	4.808	17621666	11130907	500.000	500.000
22) L5 AR-1248-2	5.911	5.049	26964885	17206728	500.000	500.000
23) L5 AR-1248-3	6.117	5.092	28653442	17917087	500.000	500.000
24) L5 AR-1248-4	6.523	5.267	23244757	20393731	500.000	500.000
25) L5 AR-1248-5	6.562	5.666	22484819	16758749	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091222\
Data File : PP051260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2022 10:19
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 11:56:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 11:55:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.457 min
Delta R.T.: 0.000 min
Response: 88749532
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.701 min
Delta R.T.: 0.000 min
Response: 63569531
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.310 min
Delta R.T.: 0.000 min
Response: 54809351
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.812 min
Delta R.T.: 0.000 min
Response: 45335771
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 17621666
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.808 min
Delta R.T.: 0.000 min
Response: 11130907
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.911 min
Delta R.T.: 0.000 min
Response: 26964885
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.049 min
Delta R.T.: 0.000 min
Response: 17206728
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.117 min
Delta R.T.: 0.000 min
Response: 28653442
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 17917087
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.523 min
Delta R.T.: 0.000 min
Response: 23244757
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 20393731
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.562 min
Delta R.T.: 0.000 min
Response: 22484819
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 16758749
Conc: 500.00 ng/ml