

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091222\
 Data File : PP051276.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2022 15:02
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP091222

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 15:33:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 14:37:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.458	3.702	85968410	62057849	54.941	53.080
2)	SA Decachlor...	10.312	8.812	48994248	45997093	46.614	52.752
Target Compounds							
21)	L5 AR-1248-1	5.635	4.808	16841971	10933528	495.055	512.833
22)	L5 AR-1248-2	5.912	5.049	26229448	16910106	497.374	510.039
23)	L5 AR-1248-3	6.118	5.092	28254210	17649603	496.287	511.115
24)	L5 AR-1248-4	6.524	5.267	22402525	20087158	497.609	510.475
25)	L5 AR-1248-5	6.563	5.667	22452817	16558056	517.314	512.760



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091222\
Data File : PP051276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2022 15:02
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP091222

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 15:33:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 14:37:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.458 min
Delta R.T.: 0.000 min
Response: 85968410
Conc: 54.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP091222




#1 Tetrachloro-m-xylene

R.T.: 3.702 min
Delta R.T.: 0.001 min
Response: 62057849
Conc: 53.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.312 min
Delta R.T.: 0.000 min
Response: 48994248
Conc: 46.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.812 min
Delta R.T.: 0.000 min
Response: 45997093
Conc: 52.75 ng/ml



#21 AR-1248-1


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 16841971
Conc: 495.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP091222




#21 AR-1248-1

R.T.: 4.808 min
Delta R.T.: 0.000 min
Response: 10933528
Conc: 512.83 ng/ml




#22 AR-1248-2

R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 26229448
Conc: 497.37 ng/ml



#22 AR-1248-2


R.T.: 5.049 min
Delta R.T.: 0.000 min
Response: 16910106
Conc: 510.04 ng/ml



#23 AR-1248-3


R.T.: 6.118 min
Delta R.T.: 0.000 min
Response: 28254210
Conc: 496.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP091222




#23 AR-1248-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 17649603
Conc: 511.11 ng/ml




#24 AR-1248-4

R.T.: 6.524 min
Delta R.T.: 0.001 min
Response: 22402525
Conc: 497.61 ng/ml



#24 AR-1248-4


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 20087158
Conc: 510.47 ng/ml



#25 AR-1248-5

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 22452817
Conc: 517.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP091222



#25 AR-1248-5

R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 16558056
Conc: 512.76 ng/ml