

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091223\
 Data File : PP059857.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 10:15
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:50:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 12 04:36:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.604	79918562	106.5E6	51.434	48.021
2) SA Decachlor...	10.155	8.619	59682834	77007212	48.845	42.219
Target Compounds						
16) L4 AR-1242-1	5.558	4.690	20338071	28454852	458.692	454.523
17) L4 AR-1242-2	5.581	4.709	29640287	41795433	463.687	454.762
18) L4 AR-1242-3	5.643	4.884	18567605	22309505	462.895	457.905
19) L4 AR-1242-4	5.742	4.969	15549188	23387258	473.939	456.149
20) L4 AR-1242-5	6.483	5.495	14078111	25732988	461.211	433.769

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091223\
Data File : PP059857.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 10:15
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:50:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 12 04:36:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

