

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091223\
 Data File : PP059859.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 11:10
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:50:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 12 04:36:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.383	3.605	85824063	115.7E6	55.235	52.189
2) SA Decachlor...	10.157	8.620	65670286	84233977	53.746	46.181
Target Compounds						
26) L6 AR-1254-1	6.419	5.495	26760317	56239281	507.149	475.558
27) L6 AR-1254-2	6.639	5.644	38747420	49079433	507.627	470.052
28) L6 AR-1254-3	7.006	6.048	38960269	74591357	504.502	470.587
29) L6 AR-1254-4	7.293	6.279	26767872	43342236	513.589	469.581
30) L6 AR-1254-5	7.715	6.697	30211132	65171359	510.577	464.513

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091123\
Data File : PP059859.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 11:10
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:50:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 12 04:36:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

