

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091223\
 Data File : PP059877.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 18:27
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:56:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 12 04:36:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.605	82339428	107.5E6	52.992	48.470
2) SA Decachlor...	10.155	8.619	64740110	81863587	52.984	44.881
Target Compounds						
21) L5 AR-1248-1	5.558	4.692	15519052	21433169	487.442	468.948
22) L5 AR-1248-2	5.833	4.929	22462526	31948371	492.519	463.608
23) L5 AR-1248-3	6.039	4.971	24630767	33705545	496.713	468.334
24) L5 AR-1248-4	6.444	5.141	25355411	38559012	515.396	467.295
25) L5 AR-1248-5	6.483	5.537	24847319	35224432	511.955	465.430



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091223\
Data File : PP059877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 18:27
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:56:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 12 04:36:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: 0.000 min
Response: 82339428
Conc: 52.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 107467794
Conc: 48.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.155 min
Delta R.T.: 0.001 min
Response: 64740110
Conc: 52.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.619 min
Delta R.T.: -0.002 min
Response: 81863587
Conc: 44.88 ng/ml



#21 AR-1248-1


R.T.: 5.558 min
Delta R.T.: 0.001 min
Response: 15519052
Conc: 487.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.692 min
Delta R.T.: 0.000 min
Response: 21433169
Conc: 468.95 ng/ml




#22 AR-1248-2

R.T.: 5.833 min
Delta R.T.: 0.000 min
Response: 22462526
Conc: 492.52 ng/ml



#22 AR-1248-2


R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 31948371
Conc: 463.61 ng/ml



#23 AR-1248-3


R.T.: 6.039 min
Delta R.T.: 0.001 min
Response: 24630767
Conc: 496.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.971 min
Delta R.T.: 0.000 min
Response: 33705545
Conc: 468.33 ng/ml




#24 AR-1248-4

R.T.: 6.444 min
Delta R.T.: 0.002 min
Response: 25355411
Conc: 515.40 ng/ml



#24 AR-1248-4


R.T.: 5.141 min
Delta R.T.: 0.000 min
Response: 38559012
Conc: 467.30 ng/ml



#25 AR-1248-5

R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 24847319
Conc: 511.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.537 min
Delta R.T.: 0.000 min
Response: 35224432
Conc: 465.43 ng/ml