

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091422\
 Data File : PP051330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 14:21
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 15:03:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 14:37:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.456	3.701	77673896	58838078	49.640	50.326
2) SA Decachlor...	10.307	8.809	67974834	46672912	64.672	53.527
Target Compounds						
16) L4 AR-1242-1	5.634	4.807	20963733	14004566	448.152	487.550
17) L4 AR-1242-2	5.657	4.826	30034231	20942294	447.663	509.332
18) L4 AR-1242-3	5.719	5.007	18997009	11211976	453.359	493.881
19) L4 AR-1242-4	5.819	5.091	15054111	11776922	454.041	487.865
20) L4 AR-1242-5	6.560	5.622	14324872	12957913	524.644	511.711

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091422\
Data File : PP051330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 14:21
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 15:03:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 14:37:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

