

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091422\
 Data File : PP051346.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 19:34
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/15/2022
 Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 23:24:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 14:37:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.457	3.702	72046077	55183200	46.043	47.200
2) SA Decachlor...	10.308	8.808	61551661	45347514	58.561	52.007
Target Compounds						
26) L6 AR-1254-1	6.496	5.623	23874057	28240518	450.530	504.670
27) L6 AR-1254-2	6.717	5.772	46041461	23950387	580.268	513.125
28) L6 AR-1254-3	7.084	6.183	50430293	37363145	587.635	507.688
29) L6 AR-1254-4	7.371	6.413	33703789	20650597	577.755	512.089
30) L6 AR-1254-5	7.792	6.836	34060690	31217926	550.403m	529.019

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091422\
Data File : PP051346.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 19:34
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 23:24:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 14:37:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.457 min
Delta R.T.: -0.001 min
Response: 72046077
Conc: 46.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022




#1 Tetrachloro-m-xylene

R.T.: 3.702 min
Delta R.T.: 0.001 min
Response: 55183200
Conc: 47.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.308 min
Delta R.T.: -0.004 min
Response: 61551661
Conc: 58.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: -0.004 min
Response: 45347514
Conc: 52.01 ng/ml




#26 AR-1254-1

R.T.: 6.496 min
Delta R.T.: 0.000 min
Response: 23874057
Conc: 450.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022




#26 AR-1254-1

R.T.: 5.623 min
Delta R.T.: -0.002 min
Response: 28240518
Conc: 504.67 ng/ml




#27 AR-1254-2

R.T.: 6.717 min
Delta R.T.: 0.000 min
Response: 46041461
Conc: 580.27 ng/ml



#27 AR-1254-2

R.T.: 5.772 min
Delta R.T.: -0.002 min
Response: 23950387
Conc: 513.12 ng/ml




#28 AR-1254-3

R.T.: 7.084 min
Delta R.T.: 0.000 min
Response: 50430293
Conc: 587.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022




#28 AR-1254-3

R.T.: 6.183 min
Delta R.T.: -0.002 min
Response: 37363145
Conc: 507.69 ng/ml




#29 AR-1254-4

R.T.: 7.371 min
Delta R.T.: 0.000 min
Response: 33703789
Conc: 577.76 ng/ml



#29 AR-1254-4

R.T.: 6.413 min
Delta R.T.: 0.000 min
Response: 20650597
Conc: 512.09 ng/ml




#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: -0.002 min
Response: 34060690
Conc: 550.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022



#30 AR-1254-5

R.T.: 6.836 min
Delta R.T.: -0.002 min
Response: 31217926
Conc: 529.02 ng/ml