

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091522\
 Data File : PP051376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2022 16:52
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 18:47:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 14:37:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.457	3.702	64449866	49408440	41.189	42.261
2) SA Decachlor...	10.306	8.808	56301620	40533525	53.566	46.486
Target Compounds						
26) L6 AR-1254-1	6.495	5.622	25436950	29248414	480.024	522.681
27) L6 AR-1254-2	6.715	5.771	39740391	23866384	500.855	511.325
28) L6 AR-1254-3	7.083	6.182	44330808	33642213	516.562	457.128
29) L6 AR-1254-4	7.370	6.412	27988681	18432556	479.786	457.086
30) L6 AR-1254-5	7.792	6.835	29843278	27783718	482.252	470.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091522\
Data File : PP051376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2022 16:52
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 18:47:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 14:37:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

