

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091522\
 Data File : PP051366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2022 14:06
 Operator : YP\AJ
 Sample : N4603-07 10X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SOM-2022-73A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 14:52:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 13 14:37:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.456	3.700	2189240	1754387	1.399	1.501
2) SA Decachlor...	10.306	8.807	2881683	1873050	2.742	2.148

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091522\
Data File : PP051366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2022 14:06
Operator : YP\AJ
Sample : N4603-07 10X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SOM-2022-73A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 14:52:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 13 14:37:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.456 min
Delta R.T.: -0.002 min
Response: 2189240
Conc: 1.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-73A




#1 Tetrachloro-m-xylene

R.T.: 3.700 min
Delta R.T.: 0.000 min
Response: 1754387
Conc: 1.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.306 min
Delta R.T.: -0.006 min
Response: 2881683
Conc: 2.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: -0.005 min
Response: 1873050
Conc: 2.15 ng/ml