

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP091823.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 18 19:32:02 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP059884.D 1000 =PP059882.D 750 =PP059883.D
 250 =PP059885.D 50 =PP059886.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.112	2.046	2.050	2.176	2.017	2.080	E6	3.07
2) SA	Decachlorobiphenyl	1.579	1.503	1.529	1.618	1.397	1.525	E6	5.54
3) L1	AR-1016-1	6.791	6.380	6.360	7.243	6.963	6.748	E4	5.64
4) L1	AR-1016-2	0.965	0.909	0.908	1.032	0.970	0.957	E5	5.39
5) L1	AR-1016-3	6.023	5.547	5.615	6.525	6.052	5.952	E4	6.62
6) L1	AR-1016-4	4.976	4.617	4.646	5.386	4.858	4.897	E4	6.36
7) L1	AR-1016-5	4.959	4.560	4.640	5.365	5.060	4.917	E4	6.65
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	8.996	8.367	8.450	9.582	9.674	9.014	E4	6.78
32) L7	AR-1260-2	1.043	0.977	0.987	1.105	1.173	1.057	E5	7.81
33) L7	AR-1260-3	8.033	7.355	7.629	8.514	7.812	7.869	E4	5.57
34) L7	AR-1260-4	9.321	8.538	8.811	9.795	9.180	9.129	E4	5.29
35) L7	AR-1260-5	1.747	1.696	1.703	1.764	1.625	1.707	E5	3.17
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP059884.D 1000 =PP059882.D 750 =PP059883.D
 250 =PP059885.D 50 =PP059886.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.670	1.639	1.633	1.704	1.499	1.629	E6	4.79
2) SA	Decachlorobiphenyl	1.564	1.517	1.529	1.591	1.370	1.514	E6	5.68
3) L1	AR-1016-1	6.098	5.766	5.774	6.425	5.994	6.012	E4	4.52
4) L1	AR-1016-2	8.412	8.189	8.057	8.737	8.107	8.300	E4	3.37
5) L1	AR-1016-3	4.569	4.356	4.352	4.799	4.353	4.486	E4	4.43

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP091823.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 18 19:32:02 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP059884.D 1000 =PP059882.D 750 =PP059883.D
 250 =PP059885.D 50 =PP059886.D

Compound			500	1000	750	250	50	Avg	%RSD

6)	L1	AR-1016-4	3.611	3.382	3.397	3.865	3.385	3.528 E4	6.00
7)	L1	AR-1016-5	4.766	4.486	4.489	5.039	4.848	4.726 E4	5.05
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	9.287	8.863	8.879	9.714	9.815	9.312 E4	4.81
32)	L7	AR-1260-2	1.103	1.057	1.060	1.146	1.116	1.096 E5	3.46
33)	L7	AR-1260-3	1.045	1.012	0.997	1.085	1.017	1.031 E5	3.38
34)	L7	AR-1260-4	8.557	8.272	8.314	8.677	7.823	8.328 E4	3.95
35)	L7	AR-1260-5	1.986	1.963	1.958	1.981	1.733	1.924 E5	5.59
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range