

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
 Data File : PP059898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2023 22:31
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:59:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 04:55:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.668	10633395	8116144	4.799	4.435
2) SA Decachlor...	10.431	8.754	7884038	7393053	4.739	4.396
Target Compounds						
21) L5 AR-1248-1	5.697	4.771	2229918	1943300	48.741	48.780
22) L5 AR-1248-2	5.975	5.012	3321548	2754225	50.534	50.137
23) L5 AR-1248-3	6.183	5.055	3611584	2859047	50.168	49.368
24) L5 AR-1248-4	6.591	5.228	4147319	3479838	53.540	51.224
25) L5 AR-1248-5	6.631	5.626	4014232	3470601	53.261	52.352



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
Data File : PP059898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2023 22:31
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:59:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 04:55:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: 0.000 min
Response: 10633395
Conc: 4.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 8116144
Conc: 4.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: 0.000 min
Response: 7884038
Conc: 4.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.754 min
Delta R.T.: 0.000 min
Response: 7393053
Conc: 4.40 ng/ml



#21 AR-1248-1


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 2229918
Conc: 48.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.771 min
Delta R.T.: -0.002 min
Response: 1943300
Conc: 48.78 ng/ml




#22 AR-1248-2

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 3321548
Conc: 50.53 ng/ml



#22 AR-1248-2


R.T.: 5.012 min
Delta R.T.: -0.001 min
Response: 2754225
Conc: 50.14 ng/ml



#23 AR-1248-3


R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 3611584
Conc: 50.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.055 min
Delta R.T.: 0.000 min
Response: 2859047
Conc: 49.37 ng/ml




#24 AR-1248-4

R.T.: 6.591 min
Delta R.T.: -0.001 min
Response: 4147319
Conc: 53.54 ng/ml



#24 AR-1248-4


R.T.: 5.228 min
Delta R.T.: 0.000 min
Response: 3479838
Conc: 51.22 ng/ml



#25 AR-1248-5

R.T.: 6.631 min
Delta R.T.: 0.000 min
Response: 4014232
Conc: 53.26 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 5.626 min
Delta R.T.: 0.000 min
Response: 3470601
Conc: 52.35 ng/ml