

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
 Data File : PP059892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2023 20:54
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:41:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 04:40:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.669	55755484	43232368	25.651	25.239
2) SA Decachlor...	10.432	8.756	42114079	41565071	25.995	26.063
Target Compounds						
16) L4 AR-1242-1	5.698	4.773	16205275	13959286	267.076	263.067
17) L4 AR-1242-2	5.721	4.793	22493069	18802881	265.414	256.987
18) L4 AR-1242-3	5.784	4.971	14350758	10494893	269.226	261.361
19) L4 AR-1242-4	5.884	5.056	11736892	10605904	266.517	265.626
20) L4 AR-1242-5	6.631	5.585	12344046	12714872	267.869	261.686



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
Data File : PP059892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2023 20:54
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:41:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 04:40:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: 0.000 min
Response: 55755484
Conc: 25.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.669 min
Delta R.T.: 0.000 min
Response: 43232368
Conc: 25.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: 0.000 min
Response: 42114079
Conc: 25.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.756 min
Delta R.T.: 0.000 min
Response: 41565071
Conc: 26.06 ng/ml



#16 AR-1242-1


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 16205275
Conc: 267.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 13959286
Conc: 263.07 ng/ml




#17 AR-1242-2

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 22493069
Conc: 265.41 ng/ml



#17 AR-1242-2


R.T.: 4.793 min
Delta R.T.: 0.000 min
Response: 18802881
Conc: 256.99 ng/ml



#18 AR-1242-3


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 14350758
Conc: 269.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.971 min
Delta R.T.: 0.000 min
Response: 10494893
Conc: 261.36 ng/ml




#19 AR-1242-4

R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 11736892
Conc: 266.52 ng/ml



#19 AR-1242-4


R.T.: 5.056 min
Delta R.T.: 0.000 min
Response: 10605904
Conc: 265.63 ng/ml



#20 AR-1242-5

R.T.: 6.631 min
Delta R.T.: 0.000 min
Response: 12344046
Conc: 267.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.585 min
Delta R.T.: 0.000 min
Response: 12714872
Conc: 261.69 ng/ml