

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
 Data File : PP059896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2023 21:59
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:58:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 04:55:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.669	110.8E6	91492094	50.000	50.000
2) SA Decachlor...	10.432	8.754	83190967	84090924	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.697	4.773	22875305	19919043	500.000	500.000
22) L5 AR-1248-2	5.976	5.013	32864598	27467179	500.000	500.000
23) L5 AR-1248-3	6.184	5.056	35994933	28956453	500.000	500.000
24) L5 AR-1248-4	6.592	5.229	38730903	33967113	500.000	500.000
25) L5 AR-1248-5	6.631	5.626	37684481	33146488	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091823\
Data File : PP059896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2023 21:59
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:58:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 04:55:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: 0.000 min
Response: 110784256
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.669 min
Delta R.T.: 0.000 min
Response: 91492094
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: 0.000 min
Response: 83190967
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.754 min
Delta R.T.: 0.000 min
Response: 84090924
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 22875305
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 19919043
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.976 min
Delta R.T.: 0.000 min
Response: 32864598
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.013 min
Delta R.T.: 0.000 min
Response: 27467179
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.184 min
Delta R.T.: 0.000 min
Response: 35994933
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.056 min
Delta R.T.: 0.000 min
Response: 28956453
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.592 min
Delta R.T.: 0.000 min
Response: 38730903
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 33967113
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.631 min
Delta R.T.: 0.000 min
Response: 37684481
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.626 min
Delta R.T.: 0.000 min
Response: 33146488
Conc: 500.00 ng/ml