

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD P\Data\PP091918\
 Data File : PP027431.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Sep 2018 16:02
 Operator : EI\MA
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

Sohil
 9/20/2018 2:44:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 06:58:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091918.M
 Quant Title : 8080.M
 QLast Update : Thu Sep 20 06:57:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.045	12.157	129.8E6	225.1E6	509.666	511.398
Target Compounds							
1) T	Dalapon	3.312	2.983	264.4E6	256.6E6	459.630	455.856
2) T	3,5-DICHL...	10.459	10.250	232.9E6	254.9E6	479.280	475.985
3) T	4-Nitroph...	11.620	11.301	200.1E6	98469137	446.559	446.199
5) T	DICAMBA	12.366	12.468	727.8E6	803.6E6	479.236	483.196
6) T	MCPD	12.707	12.657	20502936	41727785	46.310	45.868m
7) T	MCPA	12.951	13.047	31254196	66089611	45.330	45.443
8) T	DICHLORPROP	13.563	13.583	161.9E6	204.0E6	492.671	485.920
9) T	2,4-D	13.912	14.068	184.3E6	261.1E6	471.156	478.142
10) T	Pentachlo...	14.330	14.679	3943.9E6	3308.8E6	506.140	515.937
11) T	2,4,5-TP ...	15.193	15.250	1418.9E6	1250.6E6	492.796	495.564
12) T	2,4,5-T	15.581	15.889	1185.9E6	1008.0E6	479.597	479.807
13) T	2,4-DB	16.344	16.492	99146281	130.5E6	470.220	466.869
14) T	DINOSEB	17.834	16.884	882.4E6	744.7E6	468.345	468.879
15) T	Picloram	17.592	18.261	1529.5E6	1342.6E6	442.385	449.205
16) T	DCPA	18.197	18.144	2126.4E6	1616.2E6	499.366	509.379

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD P\Data\PP091918\
Data File : PP027431.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Sep 2018 16:02
Operator : EI\MA
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDICC500

Manual Integrations
APPROVED

Sohil
9/20/2018 2:44:17 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 06:58:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091918.M
Quant Title : 8080.M
QLast Update : Thu Sep 20 06:57:54 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

