

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091922\
 Data File : PP051420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2022 17:48
 Operator : YP\AJ
 Sample : PP091922ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP091922

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/20/2022
 Supervised By :Ankita Jodhani 09/20/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 18:12:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 19 18:10:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.685	93305043	79183175	52.505	52.420m
2) SA Decachlor...	10.258	8.788	78215756	62185213	50.610	51.898
Target Compounds						
3) L1 AR-1016-1	5.606	4.790	31073125	25323594	504.211	508.116
4) L1 AR-1016-2	5.628	4.809	43812852	36387748	506.624	509.712
5) L1 AR-1016-3	5.691	4.988	28000061	19476083	511.741	518.686
6) L1 AR-1016-4	5.790	5.030	21950623	15670301	515.543	515.563
7) L1 AR-1016-5	6.087	5.248	22532358	20573041	495.051	518.974
31) L7 AR-1260-1	7.219	6.297	47125554	38285165	504.803	513.111
32) L7 AR-1260-2	7.476	6.487	51625245	45005354	488.956	509.008
33) L7 AR-1260-3	7.837	6.643	38495342	42753245	498.354	504.105
34) L7 AR-1260-4	8.064	7.120	42826434	34971983	482.602	521.617
35) L7 AR-1260-5	8.389	7.362	73488338	77049814	477.674	521.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091922\
Data File : PP051420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2022 17:48
Operator : YP\AJ
Sample : PP091922ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP091922

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2022
Supervised By :Ankita Jodhani 09/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 18:12:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 19 18:10:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

