

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091922\
 Data File : PP051427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2022 19:45
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 23:14:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 19 18:10:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.687	91364890	75572849	51.413	50.030
2) SA Decachlor...	10.259	8.788	77361337	61368428	50.057	51.217
Target Compounds						
16) L4 AR-1242-1	5.607	4.791	25513594	20603788	482.102	476.419
17) L4 AR-1242-2	5.630	4.811	36205719	29299591	484.861	485.275
18) L4 AR-1242-3	5.692	4.990	23122154	15724461	488.504	491.173
19) L4 AR-1242-4	5.791	5.074	18087543	15912004	492.227	489.373
20) L4 AR-1242-5	6.532	5.607	18075641	19114780	506.923	503.348



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091922\
Data File : PP051427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2022 19:45
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 23:14:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 19 18:10:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: 0.004 min
Response: 91364890
Conc: 51.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: 0.003 min
Response: 75572849
Conc: 50.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.259 min
Delta R.T.: 0.005 min
Response: 77361337
Conc: 50.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.788 min
Delta R.T.: 0.003 min
Response: 61368428
Conc: 51.22 ng/ml



#16 AR-1242-1


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 25513594
Conc: 482.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.791 min
Delta R.T.: 0.002 min
Response: 20603788
Conc: 476.42 ng/ml




#17 AR-1242-2

R.T.: 5.630 min
Delta R.T.: 0.001 min
Response: 36205719
Conc: 484.86 ng/ml



#17 AR-1242-2


R.T.: 4.811 min
Delta R.T.: 0.002 min
Response: 29299591
Conc: 485.27 ng/ml



#18 AR-1242-3


R.T.: 5.692 min
Delta R.T.: 0.000 min
Response: 23122154
Conc: 488.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.990 min
Delta R.T.: 0.002 min
Response: 15724461
Conc: 491.17 ng/ml




#19 AR-1242-4

R.T.: 5.791 min
Delta R.T.: 0.000 min
Response: 18087543
Conc: 492.23 ng/ml



#19 AR-1242-4


R.T.: 5.074 min
Delta R.T.: 0.001 min
Response: 15912004
Conc: 489.37 ng/ml



#20 AR-1242-5

R.T.: 6.532 min
Delta R.T.: 0.002 min
Response: 18075641
Conc: 506.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.607 min
Delta R.T.: 0.002 min
Response: 19114780
Conc: 503.35 ng/ml