

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091923\
 Data File : PP059942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 13:45
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 14:08:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 09:22:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.667	101.2E6	78122186	48.630	47.961
2) SA Decachlor...	10.429	8.748	86948191	96298993	57.002	63.599
Target Compounds						
16) L4 AR-1242-1	5.697	4.771	27706567	25133668	463.897	477.272
17) L4 AR-1242-2	5.720	4.790	39734185	35094065	472.817	487.036
18) L4 AR-1242-3	5.783	4.968	24098510	19038710	460.761	486.505
19) L4 AR-1242-4	5.883	5.053	19952288	18547974	463.016	473.290
20) L4 AR-1242-5	6.630	5.582	21674620	23628971	474.078	488.404



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091923\
Data File : PP059942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 13:45
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 14:08:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 09:22:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: -0.001 min
Response: 101163461
Conc: 48.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.003 min
Response: 78122186
Conc: 47.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.429 min
Delta R.T.: -0.002 min
Response: 86948191
Conc: 57.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.748 min
Delta R.T.: -0.010 min
Response: 96298993
Conc: 63.60 ng/ml



#16 AR-1242-1


R.T.: 5.697 min
Delta R.T.: -0.002 min
Response: 27706567
Conc: 463.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.771 min
Delta R.T.: -0.003 min
Response: 25133668
Conc: 477.27 ng/ml




#17 AR-1242-2

R.T.: 5.720 min
Delta R.T.: -0.002 min
Response: 39734185
Conc: 472.82 ng/ml



#17 AR-1242-2


R.T.: 4.790 min
Delta R.T.: -0.004 min
Response: 35094065
Conc: 487.04 ng/ml



#18 AR-1242-3


R.T.: 5.783 min
Delta R.T.: -0.003 min
Response: 24098510
Conc: 460.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.968 min
Delta R.T.: -0.003 min
Response: 19038710
Conc: 486.51 ng/ml




#19 AR-1242-4

R.T.: 5.883 min
Delta R.T.: -0.002 min
Response: 19952288
Conc: 463.02 ng/ml



#19 AR-1242-4


R.T.: 5.053 min
Delta R.T.: -0.004 min
Response: 18547974
Conc: 473.29 ng/ml



#20 AR-1242-5

R.T.: 6.630 min
Delta R.T.: -0.003 min
Response: 21674620
Conc: 474.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.582 min
Delta R.T.: -0.004 min
Response: 23628971
Conc: 488.40 ng/ml